

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010423\
 Data File : PP054260.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Jan 2023 17:49
 Operator : YP\AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 05 00:12:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 05 00:08:35 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.339	3.587	49205814	40069334	25.450	24.960
2) SA Decachlor...	10.102	8.614	36819189	35759148	26.540	26.363
Target Compounds						
21) L5 AR-1248-1	5.511	4.675	12127691	8778809	266.715	261.354
22) L5 AR-1248-2	5.786	4.913	16535850	13373107	254.875	265.230
23) L5 AR-1248-3	5.990	4.955	15978146	13881268	266.798	264.374
24) L5 AR-1248-4	6.396	5.127	15541404	15732868	264.263	263.057
25) L5 AR-1248-5	6.435	5.521	14841287	13128490	264.020	257.354



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010423\
Data File : PP054260.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Jan 2023 17:49
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 05 00:12:03 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010423.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 05 00:08:35 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.339 min
Delta R.T.: 0.000 min
Response: 49205814
Conc: 25.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.587 min
Delta R.T.: 0.000 min
Response: 40069334
Conc: 24.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.102 min
Delta R.T.: 0.000 min
Response: 36819189
Conc: 26.54 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.614 min
Delta R.T.: 0.000 min
Response: 35759148
Conc: 26.36 ng/ml



#21 AR-1248-1


R.T.: 5.511 min
Delta R.T.: 0.000 min
Response: 12127691
Conc: 266.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.675 min
Delta R.T.: 0.000 min
Response: 8778809
Conc: 261.35 ng/ml




#22 AR-1248-2

R.T.: 5.786 min
Delta R.T.: 0.000 min
Response: 16535850
Conc: 254.87 ng/ml



#22 AR-1248-2


R.T.: 4.913 min
Delta R.T.: 0.000 min
Response: 13373107
Conc: 265.23 ng/ml



#23 AR-1248-3


R.T.: 5.990 min
Delta R.T.: 0.000 min
Response: 15978146
Conc: 266.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.955 min
Delta R.T.: 0.000 min
Response: 13881268
Conc: 264.37 ng/ml




#24 AR-1248-4

R.T.: 6.396 min
Delta R.T.: 0.000 min
Response: 15541404
Conc: 264.26 ng/ml



#24 AR-1248-4


R.T.: 5.127 min
Delta R.T.: 0.000 min
Response: 15732868
Conc: 263.06 ng/ml



#25 AR-1248-5

R.T.: 6.435 min
Delta R.T.: 0.000 min
Response: 14841287
Conc: 264.02 ng/ml

Instrument :
ECD_P
ClientSampleId :



#25 AR-1248-5

R.T.: 5.521 min
Delta R.T.: 0.000 min
Response: 13128490
Conc: 257.35 ng/ml