

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010423\
 Data File : PP054263.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Jan 2023 18:38
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 05 02:21:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 05 02:21:37 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.338	3.587	143.9E6	118.5E6	75.247	75.233
2) SA Decachlor...	10.103	8.613	96190994	97578388	75.035	74.983
Target Compounds						
26) L6 AR-1254-1	6.369	5.480	44203093	62199559	749.622	753.773
27) L6 AR-1254-2	6.589	5.628	66582436	53704213	751.810	753.252
28) L6 AR-1254-3	6.956	6.034	68435183	85151702	752.590	754.455
29) L6 AR-1254-4	7.243	6.263	48685759	46087053	751.188	751.880
30) L6 AR-1254-5	7.664	6.683	53570450	70375066	747.535	751.803



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010423\
Data File : PP054263.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Jan 2023 18:38
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 05 02:21:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010423.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 05 02:21:37 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.338 min
Delta R.T.: 0.000 min
Response: 143936432
Conc: 75.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.587 min
Delta R.T.: 0.000 min
Response: 118490518
Conc: 75.23 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.103 min
Delta R.T.: 0.000 min
Response: 96190994
Conc: 75.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.613 min
Delta R.T.: 0.000 min
Response: 97578388
Conc: 74.98 ng/ml



#26 AR-1254-1


R.T.: 6.369 min
Delta R.T.: 0.000 min
Response: 44203093
Conc: 749.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#26 AR-1254-1

R.T.: 5.480 min
Delta R.T.: 0.000 min
Response: 62199559
Conc: 753.77 ng/ml




#27 AR-1254-2

R.T.: 6.589 min
Delta R.T.: 0.000 min
Response: 66582436
Conc: 751.81 ng/ml



#27 AR-1254-2


R.T.: 5.628 min
Delta R.T.: 0.000 min
Response: 53704213
Conc: 753.25 ng/ml



#28 AR-1254-3


R.T.: 6.956 min
Delta R.T.: 0.000 min
Response: 68435183
Conc: 752.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#28 AR-1254-3

R.T.: 6.034 min
Delta R.T.: 0.000 min
Response: 85151702
Conc: 754.45 ng/ml




#29 AR-1254-4

R.T.: 7.243 min
Delta R.T.: 0.000 min
Response: 48685759
Conc: 751.19 ng/ml



#29 AR-1254-4


R.T.: 6.263 min
Delta R.T.: 0.000 min
Response: 46087053
Conc: 751.88 ng/ml



#30 AR-1254-5

R.T.: 7.664 min
Delta R.T.: 0.000 min
Response: 53570450
Conc: 747.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750



#30 AR-1254-5

R.T.: 6.683 min
Delta R.T.: 0.000 min
Response: 70375066
Conc: 751.80 ng/ml