

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010522\
 Data File : PP042593.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Jan 2022 6:28
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 06 15:16:59 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 05 12:01:35 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.880	4.049	972393	1295326	50.192	49.483
2)	SA Decachlor...	10.818	9.389	639563	967778	54.516	49.244
Target Compounds							
26)	L6 AR-1254-1	7.091	6.192	292097	729027	479.476	473.445
27)	L6 AR-1254-2	7.320	6.351	464785	625879	506.119	465.661
28)	L6 AR-1254-3	7.698	6.776	490558	1015063	535.507	485.995
29)	L6 AR-1254-4	7.992	7.016	339720	588894	567.165	481.188
30)	L6 AR-1254-5	8.420	7.447	353058	795502	558.418	484.344



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010522\
Data File : PP042593.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 Jan 2022 6:28
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 06 15:16:59 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010422.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 05 12:01:35 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.880 min
Delta R.T.: 0.000 min
Response: 972393
Conc: 50.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.049 min
Delta R.T.: 0.001 min
Response: 1295326
Conc: 49.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.818 min
Delta R.T.: 0.002 min
Response: 639563
Conc: 54.52 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.389 min
Delta R.T.: 0.002 min
Response: 967778
Conc: 49.24 ng/ml



#26 AR-1254-1


R.T.: 7.091 min
Delta R.T.: 0.003 min
Response: 292097
Conc: 479.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.192 min
Delta R.T.: 0.001 min
Response: 729027
Conc: 473.44 ng/ml




#27 AR-1254-2

R.T.: 7.320 min
Delta R.T.: 0.002 min
Response: 464785
Conc: 506.12 ng/ml



#27 AR-1254-2


R.T.: 6.351 min
Delta R.T.: 0.001 min
Response: 625879
Conc: 465.66 ng/ml



#28 AR-1254-3


R.T.: 7.698 min
Delta R.T.: 0.003 min
Response: 490558
Conc: 535.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.776 min
Delta R.T.: 0.002 min
Response: 1015063
Conc: 486.00 ng/ml




#29 AR-1254-4

R.T.: 7.992 min
Delta R.T.: 0.003 min
Response: 339720
Conc: 567.17 ng/ml



#29 AR-1254-4


R.T.: 7.016 min
Delta R.T.: 0.002 min
Response: 588894
Conc: 481.19 ng/ml



#30 AR-1254-5

R.T.: 8.420 min
Delta R.T.: 0.000 min
Response: 353058
Conc: 558.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.447 min
Delta R.T.: 0.002 min
Response: 795502
Conc: 484.34 ng/ml