

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010524\
 Data File : PP062720.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jan 2024 09:04
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 05 23:24:31 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 22 03:29:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.397	3.683	125.2E6	167.9E6	57.598	56.247
2)	SA Decachlor...	10.154	8.784	84339873	117.9E6	58.488	46.327
Target Compounds							
3)	L1 AR-1016-1	5.570	4.790	40404588	53163892	592.858	566.180
4)	L1 AR-1016-2	5.592	4.810	57693294	75228324	568.527	561.697
5)	L1 AR-1016-3	5.655	4.989	35360505	40629863	561.140	556.225
6)	L1 AR-1016-4	5.753	5.032	29729518	32179292	579.975	528.243
7)	L1 AR-1016-5	6.050	5.249	29955202	42796045	565.438	531.918
31)	L7 AR-1260-1	7.181	6.298	48411595	79113964	488.998	494.923
32)	L7 AR-1260-2	7.439	6.488	56211157	94527731	495.423	516.299
33)	L7 AR-1260-3	7.800	6.644	39099398	88008468	456.327	489.318
34)	L7 AR-1260-4	8.026	7.122	43496369	70552500	449.158	479.489
35)	L7 AR-1260-5	8.345	7.366	86178947	163.0E6	483.499	505.871



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010524\
Data File : PP062720.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jan 2024 09:04
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 05 23:24:31 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP122123.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 22 03:29:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.397 min
Delta R.T.: 0.001 min
Response: 125223290
Conc: 57.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.683 min
Delta R.T.: 0.000 min
Response: 167854575
Conc: 56.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.154 min
Delta R.T.: 0.012 min
Response: 84339873
Conc: 58.49 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.784 min
Delta R.T.: 0.002 min
Response: 117934932
Conc: 46.33 ng/ml



#3 AR-1016-1


R.T.: 5.570 min
Delta R.T.: 0.003 min
Response: 40404588
Conc: 592.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.790 min
Delta R.T.: 0.000 min
Response: 53163892
Conc: 566.18 ng/ml




#4 AR-1016-2

R.T.: 5.592 min
Delta R.T.: 0.002 min
Response: 57693294
Conc: 568.53 ng/ml



#4 AR-1016-2


R.T.: 4.810 min
Delta R.T.: 0.000 min
Response: 75228324
Conc: 561.70 ng/ml



#5 AR-1016-3


R.T.: 5.655 min
Delta R.T.: 0.003 min
Response: 35360505
Conc: 561.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.989 min
Delta R.T.: 0.000 min
Response: 40629863
Conc: 556.23 ng/ml




#6 AR-1016-4

R.T.: 5.753 min
Delta R.T.: 0.003 min
Response: 29729518
Conc: 579.97 ng/ml



#6 AR-1016-4


R.T.: 5.032 min
Delta R.T.: 0.000 min
Response: 32179292
Conc: 528.24 ng/ml



#7 AR-1016-5


R.T.: 6.050 min
Delta R.T.: 0.004 min
Response: 29955202
Conc: 565.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.249 min
Delta R.T.: 0.001 min
Response: 42796045
Conc: 531.92 ng/ml




#31 AR-1260-1

R.T.: 7.181 min
Delta R.T.: 0.005 min
Response: 48411595
Conc: 489.00 ng/ml



#31 AR-1260-1


R.T.: 6.298 min
Delta R.T.: 0.000 min
Response: 79113964
Conc: 494.92 ng/ml



#32 AR-1260-2


R.T.: 7.439 min
Delta R.T.: 0.004 min
Response: 56211157
Conc: 495.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.488 min
Delta R.T.: 0.001 min
Response: 94527731
Conc: 516.30 ng/ml




#33 AR-1260-3

R.T.: 7.800 min
Delta R.T.: 0.006 min
Response: 39099398
Conc: 456.33 ng/ml



#33 AR-1260-3


R.T.: 6.644 min
Delta R.T.: 0.000 min
Response: 88008468
Conc: 489.32 ng/ml



#34 AR-1260-4


R.T.: 8.026 min
Delta R.T.: 0.006 min
Response: 43496369
Conc: 449.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.122 min
Delta R.T.: 0.002 min
Response: 70552500
Conc: 479.49 ng/ml



#35 AR-1260-5

R.T.: 8.345 min
Delta R.T.: 0.007 min
Response: 86178947
Conc: 483.50 ng/ml



#35 AR-1260-5

R.T.: 7.366 min
Delta R.T.: 0.001 min
Response: 162951443
Conc: 505.87 ng/ml