

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010622\
 Data File : PP042637.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Jan 2022 20:12
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 07 03:36:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 05 12:01:35 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.883	4.050	952337	1259178	49.157	48.102
2) SA Decachlor...	10.820	9.389	665959	935080	56.766	47.580
Target Compounds						
3) L1 AR-1016-1	6.187	5.316	306882	509573	514.261	487.655
4) L1 AR-1016-2	6.210	5.336	443897	693776	509.081	490.399
5) L1 AR-1016-3	6.277	5.530	268849	397302	504.849	487.474
6) L1 AR-1016-4	6.381	5.581	227299	310163	507.523	480.402
7) L1 AR-1016-5	6.696	5.813	224598	384578	519.960	468.295
31) L7 AR-1260-1	7.869	6.913	353847	619311	582.202	475.965
32) L7 AR-1260-2	8.133	7.109	415870	726562	591.235	477.913
33) L7 AR-1260-3	8.499	7.267	316545	670744	571.350	469.155
34) L7 AR-1260-4	8.728	7.752	363678	536109	551.256	487.035
35) L7 AR-1260-5	9.046	7.999	680547	1172413	564.165	485.726



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010622\
Data File : PP042637.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 Jan 2022 20:12
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 07 03:36:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010422.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 05 12:01:35 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.883 min
Delta R.T.: 0.003 min
Response: 952337
Conc: 49.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.050 min
Delta R.T.: 0.003 min
Response: 1259178
Conc: 48.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.820 min
Delta R.T.: 0.004 min
Response: 665959
Conc: 56.77 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.389 min
Delta R.T.: 0.002 min
Response: 935080
Conc: 47.58 ng/ml



#3 AR-1016-1


R.T.: 6.187 min
Delta R.T.: 0.004 min
Response: 306882
Conc: 514.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.316 min
Delta R.T.: 0.002 min
Response: 509573
Conc: 487.66 ng/ml




#4 AR-1016-2

R.T.: 6.210 min
Delta R.T.: 0.003 min
Response: 443897
Conc: 509.08 ng/ml



#4 AR-1016-2


R.T.: 5.336 min
Delta R.T.: 0.002 min
Response: 693776
Conc: 490.40 ng/ml



#5 AR-1016-3


R.T.: 6.277 min
Delta R.T.: 0.002 min
Response: 268849
Conc: 504.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.530 min
Delta R.T.: 0.002 min
Response: 397302
Conc: 487.47 ng/ml




#6 AR-1016-4

R.T.: 6.381 min
Delta R.T.: 0.003 min
Response: 227299
Conc: 507.52 ng/ml



#6 AR-1016-4


R.T.: 5.581 min
Delta R.T.: 0.002 min
Response: 310163
Conc: 480.40 ng/ml



#7 AR-1016-5


R.T.: 6.696 min
Delta R.T.: 0.003 min
Response: 224598
Conc: 519.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.813 min
Delta R.T.: 0.002 min
Response: 384578
Conc: 468.30 ng/ml




#31 AR-1260-1

R.T.: 7.869 min
Delta R.T.: 0.002 min
Response: 353847
Conc: 582.20 ng/ml



#31 AR-1260-1


R.T.: 6.913 min
Delta R.T.: 0.002 min
Response: 619311
Conc: 475.97 ng/ml



#32 AR-1260-2


R.T.: 8.133 min
Delta R.T.: 0.003 min
Response: 415870
Conc: 591.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 7.109 min
Delta R.T.: 0.002 min
Response: 726562
Conc: 477.91 ng/ml




#33 AR-1260-3

R.T.: 8.499 min
Delta R.T.: 0.003 min
Response: 316545
Conc: 571.35 ng/ml



#33 AR-1260-3


R.T.: 7.267 min
Delta R.T.: 0.002 min
Response: 670744
Conc: 469.16 ng/ml



#34 AR-1260-4


R.T.: 8.728 min
Delta R.T.: 0.004 min
Response: 363678
Conc: 551.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.752 min
Delta R.T.: 0.002 min
Response: 536109
Conc: 487.04 ng/ml



#35 AR-1260-5

R.T.: 9.046 min
Delta R.T.: 0.002 min
Response: 680547
Conc: 564.16 ng/ml



#35 AR-1260-5

R.T.: 7.999 min
Delta R.T.: 0.001 min
Response: 1172413
Conc: 485.73 ng/ml