

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010622\
 Data File : PP042638.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Jan 2022 20:29
 Operator : AJ\MA
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 07 03:37:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 05 12:01:35 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.886	4.052	949887	1259096	49.031	48.099
2)	SA Decachlor...	10.825	9.393	639329	928298	54.496	47.235
Target Compounds							
16)	L4 AR-1242-1	6.189	5.317	234117	401691	507.607	484.923
17)	L4 AR-1242-2	6.213	5.338	340094	545415	509.376	487.535
18)	L4 AR-1242-3	6.278	5.532	203117	310719	509.531	480.471
19)	L4 AR-1242-4	6.383	5.628	170388	297779	510.288	475.209
20)	L4 AR-1242-5	7.162	6.195	169225	357963	513.503	488.036



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010622\
Data File : PP042638.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 Jan 2022 20:29
Operator : AJ\MA
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 07 03:37:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010422.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 05 12:01:35 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.886 min
Delta R.T.: 0.006 min
Response: 949887
Conc: 49.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.052 min
Delta R.T.: 0.004 min
Response: 1259096
Conc: 48.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.825 min
Delta R.T.: 0.009 min
Response: 639329
Conc: 54.50 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.393 min
Delta R.T.: 0.006 min
Response: 928298
Conc: 47.23 ng/ml



#16 AR-1242-1


R.T.: 6.189 min
Delta R.T.: 0.004 min
Response: 234117
Conc: 507.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.317 min
Delta R.T.: 0.002 min
Response: 401691
Conc: 484.92 ng/ml




#17 AR-1242-2

R.T.: 6.213 min
Delta R.T.: 0.004 min
Response: 340094
Conc: 509.38 ng/ml



#17 AR-1242-2


R.T.: 5.338 min
Delta R.T.: 0.002 min
Response: 545415
Conc: 487.54 ng/ml



#18 AR-1242-3


R.T.: 6.278 min
Delta R.T.: 0.004 min
Response: 203117
Conc: 509.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.532 min
Delta R.T.: 0.000 min
Response: 310719
Conc: 480.47 ng/ml




#19 AR-1242-4

R.T.: 6.383 min
Delta R.T.: 0.004 min
Response: 170388
Conc: 510.29 ng/ml



#19 AR-1242-4


R.T.: 5.628 min
Delta R.T.: 0.002 min
Response: 297779
Conc: 475.21 ng/ml



#20 AR-1242-5

R.T.: 7.162 min
Delta R.T.: 0.004 min
Response: 169225
Conc: 513.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 6.195 min
Delta R.T.: 0.001 min
Response: 357963
Conc: 488.04 ng/ml