

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010622\
 Data File : PP042644.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Jan 2022 22:08
 Operator : AJ\MA
 Sample : N1042-01
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TRE-1268

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/07/2022
 Supervised By :mohammad ahmed 01/10/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 07 03:40:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 05 12:01:35 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.883	4.049	374775	745879	19.345	28.493m#
2) SA Decachlor...	10.823	9.389	375594	462028	32.016m	23.509 #
Target Compounds						
26) L6 AR-1254-1	7.095	6.193	213722	345402	350.823m	224.311m#
27) L6 AR-1254-2	7.322	6.351	286303	532209	311.764m	395.969m#
28) L6 AR-1254-3	7.700	6.777	248583	576824	271.360m	276.174m
29) L6 AR-1254-4	7.995	7.016	151859	355745	253.529m	290.681
30) L6 AR-1254-5	8.422	7.446	180485	389497	285.466m	237.147m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010622\
Data File : PP042644.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 Jan 2022 22:08
Operator : AJ\MA
Sample : N1042-01
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
TRE-1268


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/07/2022
Supervised By :mohammad ahmed 01/10/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 07 03:40:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010422.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 05 12:01:35 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.883 min
Delta R.T.: 0.003 min
Response: 374775
Conc: 19.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-1268


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/07/2022
Supervised By :mohammad ahmed 01/10/2022




#1 Tetrachloro-m-xylene

R.T.: 4.049 min
Delta R.T.: 0.002 min
Response: 745879
Conc: 28.49 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.823 min
Delta R.T.: 0.006 min
Response: 375594
Conc: 32.02 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.389 min
Delta R.T.: 0.002 min
Response: 462028
Conc: 23.51 ng/ml




#26 AR-1254-1

R.T.: 7.095 min
Delta R.T.: 0.006 min
Response: 213722
Conc: 350.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-1268


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/07/2022
Supervised By :mohammad ahmed 01/10/2022




#26 AR-1254-1

R.T.: 6.193 min
Delta R.T.: 0.002 min
Response: 345402
Conc: 224.31 ng/ml m




#27 AR-1254-2

R.T.: 7.322 min
Delta R.T.: 0.004 min
Response: 286303
Conc: 311.76 ng/ml m



#27 AR-1254-2

R.T.: 6.351 min
Delta R.T.: 0.000 min
Response: 532209
Conc: 395.97 ng/ml m




#28 AR-1254-3

R.T.: 7.700 min
Delta R.T.: 0.006 min
Response: 248583
Conc: 271.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-1268


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/07/2022
Supervised By :mohammad ahmed 01/10/2022




#28 AR-1254-3

R.T.: 6.777 min
Delta R.T.: 0.003 min
Response: 576824
Conc: 276.17 ng/ml m




#29 AR-1254-4

R.T.: 7.995 min
Delta R.T.: 0.007 min
Response: 151859
Conc: 253.53 ng/ml m



#29 AR-1254-4

R.T.: 7.016 min
Delta R.T.: 0.002 min
Response: 355745
Conc: 290.68 ng/ml




#30 AR-1254-5

R.T.: 8.422 min
Delta R.T.: 0.003 min
Response: 180485
Conc: 285.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
TRE-1268

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/07/2022
Supervised By :mohammad ahmed 01/10/2022



#30 AR-1254-5

R.T.: 7.446 min
Delta R.T.: 0.001 min
Response: 389497
Conc: 237.15 ng/ml m