

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010625\
 Data File : PP068916.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jan 2025 20:13
 Operator : YP\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 07 02:41:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 07 02:39:00 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.672	3.983	103.7E6	72050731	70.596	71.839
2) SA Decachlor...	10.525	9.112	79231899	80875884	70.661	68.430
Target Compounds						
3) L1 AR-1016-1	5.834	5.085	31884667	23903404	719.911	682.906
4) L1 AR-1016-2	5.855	5.104	50152549	35171354	702.851	688.555
5) L1 AR-1016-3	5.919	5.283	30893233	19663511	712.859	690.739
6) L1 AR-1016-4	6.016	5.324	25492459	17489657	720.205	681.935
7) L1 AR-1016-5	6.311	5.541	25073462	21731554	722.068	676.829
31) L7 AR-1260-1	7.434	6.584	43644627	39759381	693.623	677.329
32) L7 AR-1260-2	7.688	6.771	53613697	44864504	703.338	667.735
33) L7 AR-1260-3	8.048	6.928	46074851	44199980	701.271	668.012
34) L7 AR-1260-4	8.281	7.402	46776949	37541946	704.686	656.190
35) L7 AR-1260-5	8.612	7.642	89910964	82599940	713.069	672.208



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010625\
Data File : PP068916.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jan 2025 20:13
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 07 02:41:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 07 02:39:00 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.672 min
Delta R.T.: 0.000 min
Response: 103656527
Conc: 70.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 72050731
Conc: 71.84 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.525 min
Delta R.T.: 0.000 min
Response: 79231899
Conc: 70.66 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.112 min
Delta R.T.: 0.000 min
Response: 80875884
Conc: 68.43 ng/ml



#3 AR-1016-1


R.T.: 5.834 min
Delta R.T.: 0.001 min
Response: 31884667
Conc: 719.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#3 AR-1016-1

R.T.: 5.085 min
Delta R.T.: 0.000 min
Response: 23903404
Conc: 682.91 ng/ml




#4 AR-1016-2

R.T.: 5.855 min
Delta R.T.: 0.001 min
Response: 50152549
Conc: 702.85 ng/ml



#4 AR-1016-2


R.T.: 5.104 min
Delta R.T.: 0.000 min
Response: 35171354
Conc: 688.56 ng/ml



#5 AR-1016-3


R.T.: 5.919 min
Delta R.T.: 0.000 min
Response: 30893233
Conc: 712.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#5 AR-1016-3

R.T.: 5.283 min
Delta R.T.: 0.000 min
Response: 19663511
Conc: 690.74 ng/ml




#6 AR-1016-4

R.T.: 6.016 min
Delta R.T.: 0.000 min
Response: 25492459
Conc: 720.21 ng/ml



#6 AR-1016-4


R.T.: 5.324 min
Delta R.T.: 0.000 min
Response: 17489657
Conc: 681.94 ng/ml



#7 AR-1016-5


R.T.: 6.311 min
Delta R.T.: 0.000 min
Response: 25073462
Conc: 722.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#7 AR-1016-5

R.T.: 5.541 min
Delta R.T.: 0.000 min
Response: 21731554
Conc: 676.83 ng/ml




#31 AR-1260-1

R.T.: 7.434 min
Delta R.T.: 0.000 min
Response: 43644627
Conc: 693.62 ng/ml



#31 AR-1260-1


R.T.: 6.584 min
Delta R.T.: 0.000 min
Response: 39759381
Conc: 677.33 ng/ml



#32 AR-1260-2


R.T.: 7.688 min
Delta R.T.: 0.000 min
Response: 53613697
Conc: 703.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#32 AR-1260-2

R.T.: 6.771 min
Delta R.T.: 0.000 min
Response: 44864504
Conc: 667.73 ng/ml




#33 AR-1260-3

R.T.: 8.048 min
Delta R.T.: 0.001 min
Response: 46074851
Conc: 701.27 ng/ml



#33 AR-1260-3


R.T.: 6.928 min
Delta R.T.: 0.000 min
Response: 44199980
Conc: 668.01 ng/ml



#34 AR-1260-4


R.T.: 8.281 min
Delta R.T.: 0.001 min
Response: 46776949
Conc: 704.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#34 AR-1260-4

R.T.: 7.402 min
Delta R.T.: 0.000 min
Response: 37541946
Conc: 656.19 ng/ml



#35 AR-1260-5

R.T.: 8.612 min
Delta R.T.: 0.000 min
Response: 89910964
Conc: 713.07 ng/ml



#35 AR-1260-5

R.T.: 7.642 min
Delta R.T.: 0.000 min
Response: 82599940
Conc: 672.21 ng/ml