

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010625\
 Data File : PP068925.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jan 2025 22:40
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 07 10:53:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 07 10:53:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.673	3.983	38748606	26862031	26.450	26.455
2)	SA Decachlor...	10.525	9.111	31164692	32201952	27.361	27.308
Target Compounds							
16)	L4 AR-1242-1	5.835	5.085	9922002	7913679	267.582	276.041
17)	L4 AR-1242-2	5.856	5.104	15178283	11391184	264.152	272.365
18)	L4 AR-1242-3	5.920	5.284	9148480	6072135	261.524	264.145
19)	L4 AR-1242-4	6.017	5.368	7633794	6421018	265.004	267.328
20)	L4 AR-1242-5	6.752	5.896	8518390	7628950	273.973	267.142


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010625\
Data File : PP068925.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jan 2025 22:40
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 07 10:53:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 07 10:53:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.673 min
Delta R.T.: 0.000 min
Response: 38748606
Conc: 26.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 26862031
Conc: 26.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.525 min
Delta R.T.: 0.000 min
Response: 31164692
Conc: 27.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.111 min
Delta R.T.: 0.000 min
Response: 32201952
Conc: 27.31 ng/ml



#16 AR-1242-1


R.T.: 5.835 min
Delta R.T.: 0.000 min
Response: 9922002
Conc: 267.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 5.085 min
Delta R.T.: 0.000 min
Response: 7913679
Conc: 276.04 ng/ml




#17 AR-1242-2

R.T.: 5.856 min
Delta R.T.: 0.000 min
Response: 15178283
Conc: 264.15 ng/ml



#17 AR-1242-2


R.T.: 5.104 min
Delta R.T.: 0.000 min
Response: 11391184
Conc: 272.36 ng/ml



#18 AR-1242-3


R.T.: 5.920 min
Delta R.T.: 0.000 min
Response: 9148480
Conc: 261.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 5.284 min
Delta R.T.: 0.000 min
Response: 6072135
Conc: 264.15 ng/ml




#19 AR-1242-4

R.T.: 6.017 min
Delta R.T.: 0.000 min
Response: 7633794
Conc: 265.00 ng/ml



#19 AR-1242-4


R.T.: 5.368 min
Delta R.T.: 0.000 min
Response: 6421018
Conc: 267.33 ng/ml



#20 AR-1242-5

R.T.: 6.752 min
Delta R.T.: 0.000 min
Response: 8518390
Conc: 273.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.896 min
Delta R.T.: 0.000 min
Response: 7628950
Conc: 267.14 ng/ml