

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010722\
 Data File : PP042685.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Jan 2022 9:55
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 08 01:49:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 05 12:01:35 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.884	4.051	978159	1239184	50.490	47.338
2) SA Decachlor...	10.823	9.390	635057	935798	54.132	47.616
Target Compounds						
3) L1 AR-1016-1	6.187	5.316	307872	512140	515.918	490.112
4) L1 AR-1016-2	6.211	5.337	456110	692611	523.087	489.576
5) L1 AR-1016-3	6.277	5.531	271610	396899	510.032	486.979
6) L1 AR-1016-4	6.382	5.581	224711	313452	501.746	485.497
7) L1 AR-1016-5	6.697	5.813	219752	398696	508.741	485.486
31) L7 AR-1260-1	7.870	6.914	346286	624464	569.762	479.925
32) L7 AR-1260-2	8.135	7.111	404760	721565	575.441	474.626
33) L7 AR-1260-3	8.502	7.268	305873	668793	552.088	467.791
34) L7 AR-1260-4	8.730	7.753	361208	537704	547.513	488.484
35) L7 AR-1260-5	9.047	8.001	662045	1168305	548.827	484.024



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010722\
Data File : PP042685.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Jan 2022 9:55
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 08 01:49:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010422.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 05 12:01:35 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.884 min
Delta R.T.: 0.004 min
Response: 978159
Conc: 50.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.051 min
Delta R.T.: 0.003 min
Response: 1239184
Conc: 47.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.823 min
Delta R.T.: 0.006 min
Response: 635057
Conc: 54.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.390 min
Delta R.T.: 0.003 min
Response: 935798
Conc: 47.62 ng/ml



#3 AR-1016-1


R.T.: 6.187 min
Delta R.T.: 0.004 min
Response: 307872
Conc: 515.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.316 min
Delta R.T.: 0.003 min
Response: 512140
Conc: 490.11 ng/ml




#4 AR-1016-2

R.T.: 6.211 min
Delta R.T.: 0.004 min
Response: 456110
Conc: 523.09 ng/ml



#4 AR-1016-2


R.T.: 5.337 min
Delta R.T.: 0.002 min
Response: 692611
Conc: 489.58 ng/ml



#5 AR-1016-3


R.T.: 6.277 min
Delta R.T.: 0.003 min
Response: 271610
Conc: 510.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.531 min
Delta R.T.: 0.003 min
Response: 396899
Conc: 486.98 ng/ml




#6 AR-1016-4

R.T.: 6.382 min
Delta R.T.: 0.004 min
Response: 224711
Conc: 501.75 ng/ml



#6 AR-1016-4


R.T.: 5.581 min
Delta R.T.: 0.002 min
Response: 313452
Conc: 485.50 ng/ml



#7 AR-1016-5


R.T.: 6.697 min
Delta R.T.: 0.004 min
Response: 219752
Conc: 508.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.813 min
Delta R.T.: 0.002 min
Response: 398696
Conc: 485.49 ng/ml




#31 AR-1260-1

R.T.: 7.870 min
Delta R.T.: 0.003 min
Response: 346286
Conc: 569.76 ng/ml



#31 AR-1260-1


R.T.: 6.914 min
Delta R.T.: 0.003 min
Response: 624464
Conc: 479.93 ng/ml



#32 AR-1260-2


R.T.: 8.135 min
Delta R.T.: 0.005 min
Response: 404760
Conc: 575.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 7.111 min
Delta R.T.: 0.003 min
Response: 721565
Conc: 474.63 ng/ml




#33 AR-1260-3

R.T.: 8.502 min
Delta R.T.: 0.005 min
Response: 305873
Conc: 552.09 ng/ml



#33 AR-1260-3


R.T.: 7.268 min
Delta R.T.: 0.003 min
Response: 668793
Conc: 467.79 ng/ml



#34 AR-1260-4


R.T.: 8.730 min
Delta R.T.: 0.005 min
Response: 361208
Conc: 547.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.753 min
Delta R.T.: 0.003 min
Response: 537704
Conc: 488.48 ng/ml



#35 AR-1260-5

R.T.: 9.047 min
Delta R.T.: 0.003 min
Response: 662045
Conc: 548.83 ng/ml



#35 AR-1260-5

R.T.: 8.001 min
Delta R.T.: 0.003 min
Response: 1168305
Conc: 484.02 ng/ml