

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010722\
 Data File : PP042735.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Jan 2022 00:59
 Operator : AJ\MA
 Sample : N1065-14
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AMI-4

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/10/2022
 Supervised By :mohammad ahmed 01/10/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 08 02:11:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 05 12:01:35 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.882	4.050	449356	600308	23.195	22.932
2) SA Decachlor...	10.817	9.386	287499	397189	24.506	20.210
Target Compounds						
26) L6 AR-1254-1	7.092	6.194	6746	15190	11.074m	9.865
27) L6 AR-1254-2	7.319	6.349	11743	13100	12.787	9.746
28) L6 AR-1254-3	7.696	6.774	11944	16887	13.039	8.085m#
29) L6 AR-1254-4	7.991	7.012	9956	9089	16.622	7.426m#
30) L6 AR-1254-5	8.419	7.444	12325	28274	19.494	17.215m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010722\
Data File : PP042735.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Jan 2022 00:59
Operator : AJ\MA
Sample : N1065-14
Misc :
ALS Vial : 37 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AMI-4


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/10/2022
Supervised By :mohammad ahmed 01/10/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 08 02:11:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010422.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 05 12:01:35 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.882 min
Delta R.T.: 0.002 min
Response: 449356
Conc: 23.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AMI-4


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/10/2022
Supervised By :mohammad ahmed 01/10/2022




#1 Tetrachloro-m-xylene

R.T.: 4.050 min
Delta R.T.: 0.002 min
Response: 600308
Conc: 22.93 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.817 min
Delta R.T.: 0.000 min
Response: 287499
Conc: 24.51 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.386 min
Delta R.T.: 0.000 min
Response: 397189
Conc: 20.21 ng/ml




#26 AR-1254-1

R.T.: 7.092 min
Delta R.T.: 0.003 min
Response: 6746
Conc: 11.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AMI-4


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/10/2022
Supervised By :mohammad ahmed 01/10/2022




#26 AR-1254-1

R.T.: 6.194 min
Delta R.T.: 0.002 min
Response: 15190
Conc: 9.86 ng/ml




#27 AR-1254-2

R.T.: 7.319 min
Delta R.T.: 0.002 min
Response: 11743
Conc: 12.79 ng/ml



#27 AR-1254-2

R.T.: 6.349 min
Delta R.T.: -0.001 min
Response: 13100
Conc: 9.75 ng/ml




#28 AR-1254-3

R.T.: 7.696 min
Delta R.T.: 0.002 min
Response: 11944
Conc: 13.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AMI-4


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/10/2022
Supervised By :mohammad ahmed 01/10/2022




#28 AR-1254-3

R.T.: 6.774 min
Delta R.T.: 0.000 min
Response: 16887
Conc: 8.09 ng/ml m




#29 AR-1254-4

R.T.: 7.991 min
Delta R.T.: 0.003 min
Response: 9956
Conc: 16.62 ng/ml



#29 AR-1254-4

R.T.: 7.012 min
Delta R.T.: -0.003 min
Response: 9089
Conc: 7.43 ng/ml m




#30 AR-1254-5

R.T.: 8.419 min
Delta R.T.: 0.000 min
Response: 12325
Conc: 19.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AMI-4

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/10/2022
Supervised By :mohammad ahmed 01/10/2022



#30 AR-1254-5

R.T.: 7.444 min
Delta R.T.: 0.000 min
Response: 28274
Conc: 17.21 ng/ml m