

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010722\
 Data File : PP042748.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Jan 2022 5:01
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/10/2022
 Supervised By :mohammad ahmed 01/10/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 08 06:08:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 05 12:01:35 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.881	4.049	962307	1259776	49.672	48.125
2) SA Decachlor...	10.813	9.384	696625	946244	59.380	48.148
Target Compounds						
26) L6 AR-1254-1	7.090	6.190	304483	741399	499.807	481.480
27) L6 AR-1254-2	7.318	6.348	489467	671597	532.996	499.676
28) L6 AR-1254-3	7.696	6.773	526283	1024499	574.506	490.513
29) L6 AR-1254-4	7.989	7.012	347400	594667	579.987m	485.906
30) L6 AR-1254-5	8.417	7.443	375234	804521	593.494m	489.836

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010722\
Data File : PP042748.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Jan 2022 5:01
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/10/2022
Supervised By :mohammad ahmed 01/10/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 08 06:08:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010422.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 05 12:01:35 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.881 min
Delta R.T.: 0.001 min
Response: 962307
Conc: 49.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/10/2022
Supervised By :mohammad ahmed 01/10/2022




#1 Tetrachloro-m-xylene

R.T.: 4.049 min
Delta R.T.: 0.001 min
Response: 1259776
Conc: 48.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.813 min
Delta R.T.: -0.003 min
Response: 696625
Conc: 59.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.384 min
Delta R.T.: -0.003 min
Response: 946244
Conc: 48.15 ng/ml




#26 AR-1254-1

R.T.: 7.090 min
Delta R.T.: 0.000 min
Response: 304483
Conc: 499.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/10/2022
Supervised By :mohammad ahmed 01/10/2022




#26 AR-1254-1

R.T.: 6.190 min
Delta R.T.: -0.002 min
Response: 741399
Conc: 481.48 ng/ml




#27 AR-1254-2

R.T.: 7.318 min
Delta R.T.: 0.000 min
Response: 489467
Conc: 533.00 ng/ml



#27 AR-1254-2

R.T.: 6.348 min
Delta R.T.: -0.002 min
Response: 671597
Conc: 499.68 ng/ml




#28 AR-1254-3

R.T.: 7.696 min
Delta R.T.: 0.000 min
Response: 526283
Conc: 574.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/10/2022
Supervised By :mohammad ahmed 01/10/2022




#28 AR-1254-3

R.T.: 6.773 min
Delta R.T.: -0.002 min
Response: 1024499
Conc: 490.51 ng/ml




#29 AR-1254-4

R.T.: 7.989 min
Delta R.T.: 0.000 min
Response: 347400
Conc: 579.99 ng/ml m



#29 AR-1254-4

R.T.: 7.012 min
Delta R.T.: -0.002 min
Response: 594667
Conc: 485.91 ng/ml




#30 AR-1254-5

R.T.: 8.417 min
Delta R.T.: -0.002 min
Response: 375234
Conc: 593.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/10/2022
Supervised By :mohammad ahmed 01/10/2022



#30 AR-1254-5

R.T.: 7.443 min
Delta R.T.: -0.002 min
Response: 804521
Conc: 489.84 ng/ml