

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010919\
 Data File : PP028716.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Jan 2019 13:57
 Operator : EI\MA
 Sample : HSTDCCC750
 Misc : hexane
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 1/10/2019 1:43:39 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 10 11:37:48 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010319.M
 Quant Title : 8080.M
 QLast Update : Fri Jan 04 01:40:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.136	12.246	198.6E6	357.6E6	772.393	818.997
Target Compounds							
1) T	Dalapon	3.363	3.030	410.4E6	396.4E6	671.089	678.150
2) T	3,5-DICHL...	10.547	10.336	352.7E6	398.8E6	705.771	773.865
3) T	4-Nitroph...	11.714	11.389	300.4E6	147.6E6	759.479	750.405m
5) T	DICAMBA	12.458	12.556	1138.0E6	1321.9E6	726.942	765.417
6) T	MCPD	12.800	12.747	33368983	71118222	66.089	76.892
7) T	MCPA	13.044	13.137	53643441	109.3E6	67.391	74.708
8) T	DICHLORPROP	13.655	13.671	245.9E6	327.2E6	759.964	753.426
9) T	2,4-D	14.005	14.159	244.1E6	385.1E6	795.112m	789.399
10) T	Pentachlo...	14.428	14.774	5702.6E6	4924.5E6	726.336	757.380
11) T	2,4,5-TP ...	15.289	15.341	2080.9E6	1918.9E6	732.195	776.838
12) T	2,4,5-T	15.679	15.985	1896.0E6	1678.8E6	751.742	773.896
13) T	2,4-DB	16.442	16.584	137.0E6	190.3E6	753.132m	772.270m
14) T	DINOSEB	17.934	16.978	1488.3E6	1327.7E6	680.063	756.582
15) T	Picloram	17.699	18.361	3184.2E6	2816.7E6	745.196	759.218
16) T	DCPA	18.299	18.240	3003.9E6	2440.7E6	727.558	747.323

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010919\
Data File : PP028716.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Jan 2019 13:57
Operator : EI\MA
Sample : HSTDCCC750
Misc : hexane
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
1/10/2019 1:43:39 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 10 11:37:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010319.M
Quant Title : 8080.M
QLast Update : Fri Jan 04 01:40:06 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

