

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011024\
 Data File : PP062816.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jan 2024 00:19
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 11 02:07:34 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 11 02:00:33 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.395	3.681	194.7E6	261.1E6	75.002	74.948
2) SA Decachlor...	10.158	8.788	202.6E6	296.5E6	74.287	74.537
Target Compounds						
41) L9 AR-1268-1	8.657	7.654	161.8E6	280.1E6	743.765	746.083
42) L9 AR-1268-2	8.751	7.720	146.9E6	257.0E6	747.227	745.154
43) L9 AR-1268-3	8.979	7.929	130.6E6	224.9E6	743.997	744.708
44) L9 AR-1268-4	9.404	8.222	52089285	86115583	716.385	743.723
45) L9 AR-1268-5	9.819	8.523	402.9E6	619.1E6	745.680	747.052



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011024\
Data File : PP062816.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jan 2024 00:19
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 11 02:07:34 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 11 02:00:33 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.395 min
Delta R.T.: 0.000 min
Response: 194718035
Conc: 75.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.681 min
Delta R.T.: 0.000 min
Response: 261090467
Conc: 74.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.158 min
Delta R.T.: 0.000 min
Response: 202563525
Conc: 74.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.788 min
Delta R.T.: 0.000 min
Response: 296546353
Conc: 74.54 ng/ml



#41 AR-1268-1


R.T.: 8.657 min
Delta R.T.: 0.000 min
Response: 161834307
Conc: 743.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.654 min
Delta R.T.: 0.000 min
Response: 280052578
Conc: 746.08 ng/ml




#42 AR-1268-2

R.T.: 8.751 min
Delta R.T.: 0.000 min
Response: 146891161
Conc: 747.23 ng/ml



#42 AR-1268-2


R.T.: 7.720 min
Delta R.T.: 0.000 min
Response: 256951194
Conc: 745.15 ng/ml



#43 AR-1268-3


R.T.: 8.979 min
Delta R.T.: 0.000 min
Response: 130589129
Conc: 744.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.929 min
Delta R.T.: 0.000 min
Response: 224894099
Conc: 744.71 ng/ml




#44 AR-1268-4

R.T.: 9.404 min
Delta R.T.: 0.000 min
Response: 52089285
Conc: 716.39 ng/ml



#44 AR-1268-4


R.T.: 8.222 min
Delta R.T.: 0.000 min
Response: 86115583
Conc: 743.72 ng/ml



#45 AR-1268-5

R.T.: 9.819 min
Delta R.T.: 0.000 min
Response: 402906166
Conc: 745.68 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.523 min
Delta R.T.: 0.000 min
Response: 619116211
Conc: 747.05 ng/ml