

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011024\
 Data File : PP062792.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jan 2024 17:47
 Operator : YP\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 11 00:54:46 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 11 00:49:05 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.397	3.682	260.7E6	347.6E6	99.330	98.851
2) SA Decachlor...	10.160	8.789	151.0E6	219.2E6	97.949	96.840
Target Compounds						
3) L1 AR-1016-1	5.570	4.791	75979512	99994698	970.066	972.307
4) L1 AR-1016-2	5.593	4.810	114.3E6	142.1E6	978.870	977.684
5) L1 AR-1016-3	5.655	4.990	68740007	76858615	966.103	968.664
6) L1 AR-1016-4	5.754	5.032	57340545	61376372	968.695	956.323
7) L1 AR-1016-5	6.050	5.250	57641020	81878274	964.558	956.618
31) L7 AR-1260-1	7.183	6.300	96871089	156.3E6	969.044	968.374
32) L7 AR-1260-2	7.441	6.491	104.4E6	178.9E6	969.840	969.166
33) L7 AR-1260-3	7.803	6.647	77421097	173.6E6	958.458	972.644
34) L7 AR-1260-4	8.029	7.124	82929126	138.1E6	968.400	968.927
35) L7 AR-1260-5	8.350	7.368	154.1E6	308.3E6	982.001	984.560



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011024\
Data File : PP062792.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jan 2024 17:47
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 11 00:54:46 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 11 00:49:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.397 min
Delta R.T.: 0.000 min
Response: 260672893
Conc: 99.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.682 min
Delta R.T.: 0.000 min
Response: 347562123
Conc: 98.85 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.160 min
Delta R.T.: 0.000 min
Response: 151007649
Conc: 97.95 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.789 min
Delta R.T.: 0.000 min
Response: 219160654
Conc: 96.84 ng/ml



#3 AR-1016-1


R.T.: 5.570 min
Delta R.T.: 0.000 min
Response: 75979512
Conc: 970.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#3 AR-1016-1

R.T.: 4.791 min
Delta R.T.: 0.000 min
Response: 99994698
Conc: 972.31 ng/ml




#4 AR-1016-2

R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 114299062
Conc: 978.87 ng/ml



#4 AR-1016-2


R.T.: 4.810 min
Delta R.T.: 0.000 min
Response: 142140884
Conc: 977.68 ng/ml



#5 AR-1016-3


R.T.: 5.655 min
Delta R.T.: 0.000 min
Response: 68740007
Conc: 966.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#5 AR-1016-3

R.T.: 4.990 min
Delta R.T.: 0.000 min
Response: 76858615
Conc: 968.66 ng/ml




#6 AR-1016-4

R.T.: 5.754 min
Delta R.T.: 0.000 min
Response: 57340545
Conc: 968.69 ng/ml



#6 AR-1016-4


R.T.: 5.032 min
Delta R.T.: 0.000 min
Response: 61376372
Conc: 956.32 ng/ml



#7 AR-1016-5


R.T.: 6.050 min
Delta R.T.: 0.000 min
Response: 57641020
Conc: 964.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#7 AR-1016-5

R.T.: 5.250 min
Delta R.T.: 0.000 min
Response: 81878274
Conc: 956.62 ng/ml




#31 AR-1260-1

R.T.: 7.183 min
Delta R.T.: 0.000 min
Response: 96871089
Conc: 969.04 ng/ml



#31 AR-1260-1


R.T.: 6.300 min
Delta R.T.: 0.000 min
Response: 156260078
Conc: 968.37 ng/ml



#32 AR-1260-2


R.T.: 7.441 min
Delta R.T.: 0.000 min
Response: 104371792
Conc: 969.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#32 AR-1260-2

R.T.: 6.491 min
Delta R.T.: 0.000 min
Response: 178923747
Conc: 969.17 ng/ml




#33 AR-1260-3

R.T.: 7.803 min
Delta R.T.: 0.000 min
Response: 77421097
Conc: 958.46 ng/ml



#33 AR-1260-3


R.T.: 6.647 min
Delta R.T.: 0.000 min
Response: 173647571
Conc: 972.64 ng/ml



#34 AR-1260-4


R.T.: 8.029 min
Delta R.T.: 0.000 min
Response: 82929126
Conc: 968.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#34 AR-1260-4

R.T.: 7.124 min
Delta R.T.: 0.000 min
Response: 138081471
Conc: 968.93 ng/ml



#35 AR-1260-5

R.T.: 8.350 min
Delta R.T.: 0.000 min
Response: 154099901
Conc: 982.00 ng/ml



#35 AR-1260-5

R.T.: 7.368 min
Delta R.T.: 0.000 min
Response: 308346397
Conc: 984.56 ng/ml