

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011119\
 Data File : PP028768.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Jan 2019 17:20
 Operator : EI\MA
 Sample : HSTDCCC750
 Misc : hexane
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 11 17:41:51 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010319.M
 Quant Title : 8080.M
 QLast Update : Fri Jan 04 01:40:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.139	12.248	185.5E6	337.6E6	721.302	773.271
Target Compounds							
1) T	Dalapon	3.364	3.029	395.3E6	386.1E6	646.462	660.490
2) T	3,5-DICHL...	10.550	10.338	333.6E6	378.4E6	667.487	734.363
3) T	4-Nitroph...	11.717	11.393	290.8E6	159.1E6	735.345	808.945
5) T	DICAMBA	12.461	12.559	1045.9E6	1246.2E6	668.056	721.613
6) T	MCPD	12.804	12.750	35119060	65604634	69.339	70.931
7) T	MCPA	13.047	13.140	49964767	104.7E6	62.804	71.522
8) T	DICHLORPROP	13.658	13.675	230.0E6	319.5E6	711.038	735.841
9) T	2,4-D	14.010	14.162	227.3E6	368.3E6	740.591	755.068
10) T	Pentachlo...	14.431	14.777	5572.3E6	4853.1E6	709.734	746.394
11) T	2,4,5-TP ...	15.292	15.345	1951.4E6	1823.9E6	686.651	738.400
12) T	2,4,5-T	15.683	15.989	1797.1E6	1594.6E6	712.533	735.088
13) T	2,4-DB	16.446	16.588	138.6E6	190.0E6	761.061	771.027
14) T	DINOSEB	17.937	16.981	1460.4E6	1313.6E6	667.291	748.569
15) T	Picloram	17.703	18.365	2975.2E6	2708.6E6	697.344	729.151
16) T	DCPA	18.301	18.243	2870.5E6	2356.0E6	695.259	721.412



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011119\
Data File : PP028768.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Jan 2019 17:20
Operator : EI\MA
Sample : HSTDCCC750
Misc : hexane
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 11 17:41:51 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010319.M
Quant Title : 8080.M
QLast Update : Fri Jan 04 01:40:06 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.364 min
Delta R.T.: 0.005 min
Response: 395322253
Conc: 646.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#1 Dalapon

R.T.: 3.029 min
Delta R.T.: 0.002 min
Response: 386114422
Conc: 660.49 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 10.550 min
Delta R.T.: 0.002 min
Response: 333564466
Conc: 667.49 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 10.338 min
Delta R.T.: 0.000 min
Response: 378406665
Conc: 734.36 ng/ml



#3 4-Nitrophenol


R.T.: 11.717 min
Delta R.T.: -0.004 min
Response: 290823969
Conc: 735.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#3 4-Nitrophenol

R.T.: 11.393 min
Delta R.T.: -0.013 min
Response: 159106734
Conc: 808.95 ng/ml




#4 2,4-DCAA

R.T.: 12.139 min
Delta R.T.: 0.003 min
Response: 185450452
Conc: 721.30 ng/ml



#4 2,4-DCAA


R.T.: 12.248 min
Delta R.T.: 0.000 min
Response: 337595316
Conc: 773.27 ng/ml



#5 DICAMBA


R.T.: 12.461 min
Delta R.T.: 0.005 min
Response: 1045857083
Conc: 668.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#5 DICAMBA

R.T.: 12.559 min
Delta R.T.: 0.005 min
Response: 1246236843
Conc: 721.61 ng/ml




#6 MCPP

R.T.: 12.804 min
Delta R.T.: 0.012 min
Response: 35119060
Conc: 69.34 ug/ml



#6 MCPP


R.T.: 12.750 min
Delta R.T.: 0.009 min
Response: 65604634
Conc: 70.93 ug/ml



#7 MCPA


R.T.: 13.047 min
Delta R.T.: 0.011 min
Response: 49964767
Conc: 62.80 ug/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#7 MCPA

R.T.: 13.140 min
Delta R.T.: 0.010 min
Response: 104670839
Conc: 71.52 ug/ml




#8 DICHLORPROP

R.T.: 13.658 min
Delta R.T.: 0.003 min
Response: 230028322
Conc: 711.04 ng/ml



#8 DICHLORPROP


R.T.: 13.675 min
Delta R.T.: 0.002 min
Response: 319543845
Conc: 735.84 ng/ml



#9 2,4-D


R.T.: 14.010 min
Delta R.T.: -0.003 min
Response: 227317800
Conc: 740.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#9 2,4-D

R.T.: 14.162 min
Delta R.T.: -0.009 min
Response: 368304607
Conc: 755.07 ng/ml




#10 Pentachlorophenol

R.T.: 14.431 min
Delta R.T.: 0.006 min
Response: 5572299367
Conc: 709.73 ng/ml



#10 Pentachlorophenol


R.T.: 14.777 min
Delta R.T.: 0.004 min
Response: 4853080365
Conc: 746.39 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 15.292 min
Delta R.T.: -0.005 min
Response: 1951440568
Conc: 686.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#11 2,4,5-TP (SILVEX)

R.T.: 15.345 min
Delta R.T.: -0.004 min
Response: 1823945187
Conc: 738.40 ng/ml




#12 2,4,5-T

R.T.: 15.683 min
Delta R.T.: -0.017 min
Response: 1797092064
Conc: 712.53 ng/ml



#12 2,4,5-T


R.T.: 15.989 min
Delta R.T.: -0.013 min
Response: 1594599108
Conc: 735.09 ng/ml



#13 2,4-DB


R.T.: 16.446 min
Delta R.T.: -0.036 min
Response: 138592888
Conc: 761.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#13 2,4-DB

R.T.: 16.588 min
Delta R.T.: -0.022 min
Response: 190004900
Conc: 771.03 ng/ml




#14 DINOSEB

R.T.: 17.937 min
Delta R.T.: 0.002 min
Response: 1460356759
Conc: 667.29 ng/ml



#14 DINOSEB


R.T.: 16.981 min
Delta R.T.: 0.001 min
Response: 1313636517
Conc: 748.57 ng/ml



#15 Picloram


R.T.: 17.703 min
Delta R.T.: -0.043 min
Response: 2975166912
Conc: 697.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#15 Picloram

R.T.: 18.365 min
Delta R.T.: -0.020 min
Response: 2708615816
Conc: 729.15 ng/ml



#16 DCPA

R.T.: 18.301 min
Delta R.T.: 0.003 min
Response: 2870513320
Conc: 695.26 ng/ml



#16 DCPA

R.T.: 18.243 min
Delta R.T.: 0.003 min
Response: 2356035152
Conc: 721.41 ng/ml