

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011123\
 Data File : PP054492.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jan 2023 05:32
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 12 06:00:43 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 05 03:46:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.335	3.585	109.2E6	81393422	53.926	50.461
2) SA Decachlor...	10.093	8.603	62142883	71495534	43.269	50.682
Target Compounds						
3) L1 AR-1016-1	5.506	4.670	35594342	26710085	481.676	513.336
4) L1 AR-1016-2	5.528	4.688	49840825	37959149	467.030	529.988
5) L1 AR-1016-3	5.591	4.865	31458181	20291481	459.826	517.271
6) L1 AR-1016-4	5.689	4.907	22179439	16385275	425.999	476.124
7) L1 AR-1016-5	5.985	5.120	20253176	23742392	447.320	540.000
31) L7 AR-1260-1	7.116	6.157	35177061	41422499	451.252	488.982
32) L7 AR-1260-2	7.374	6.346	40532902	47956754	452.536	501.429
33) L7 AR-1260-3	7.735	6.499	29759240	45423188	435.484	496.945
34) L7 AR-1260-4	7.961	6.973	34599518	37524204	434.280	495.009
35) L7 AR-1260-5	8.278	7.216	62769365	79395459	437.663	501.215



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011123\
Data File : PP054492.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jan 2023 05:32
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 12 06:00:43 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010423.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 05 03:46:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.335 min
Delta R.T.: -0.003 min
Response: 109203154
Conc: 53.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.585 min
Delta R.T.: -0.001 min
Response: 81393422
Conc: 50.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.093 min
Delta R.T.: -0.008 min
Response: 62142883
Conc: 43.27 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.603 min
Delta R.T.: -0.009 min
Response: 71495534
Conc: 50.68 ng/ml



#3 AR-1016-1


R.T.: 5.506 min
Delta R.T.: -0.004 min
Response: 35594342
Conc: 481.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.670 min
Delta R.T.: -0.005 min
Response: 26710085
Conc: 513.34 ng/ml




#4 AR-1016-2

R.T.: 5.528 min
Delta R.T.: -0.005 min
Response: 49840825
Conc: 467.03 ng/ml



#4 AR-1016-2


R.T.: 4.688 min
Delta R.T.: -0.005 min
Response: 37959149
Conc: 529.99 ng/ml



#5 AR-1016-3


R.T.: 5.591 min
Delta R.T.: -0.004 min
Response: 31458181
Conc: 459.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.865 min
Delta R.T.: -0.005 min
Response: 20291481
Conc: 517.27 ng/ml




#6 AR-1016-4

R.T.: 5.689 min
Delta R.T.: -0.004 min
Response: 22179439
Conc: 426.00 ng/ml



#6 AR-1016-4


R.T.: 4.907 min
Delta R.T.: -0.006 min
Response: 16385275
Conc: 476.12 ng/ml



#7 AR-1016-5


R.T.: 5.985 min
Delta R.T.: -0.005 min
Response: 20253176
Conc: 447.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.120 min
Delta R.T.: -0.006 min
Response: 23742392
Conc: 540.00 ng/ml




#31 AR-1260-1

R.T.: 7.116 min
Delta R.T.: -0.006 min
Response: 35177061
Conc: 451.25 ng/ml



#31 AR-1260-1


R.T.: 6.157 min
Delta R.T.: -0.007 min
Response: 41422499
Conc: 488.98 ng/ml



#32 AR-1260-2


R.T.: 7.374 min
Delta R.T.: -0.006 min
Response: 40532902
Conc: 452.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.346 min
Delta R.T.: -0.007 min
Response: 47956754
Conc: 501.43 ng/ml




#33 AR-1260-3

R.T.: 7.735 min
Delta R.T.: -0.006 min
Response: 29759240
Conc: 435.48 ng/ml



#33 AR-1260-3


R.T.: 6.499 min
Delta R.T.: -0.007 min
Response: 45423188
Conc: 496.95 ng/ml



#34 AR-1260-4


R.T.: 7.961 min
Delta R.T.: -0.005 min
Response: 34599518
Conc: 434.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.973 min
Delta R.T.: -0.008 min
Response: 37524204
Conc: 495.01 ng/ml



#35 AR-1260-5

R.T.: 8.278 min
Delta R.T.: -0.005 min
Response: 62769365
Conc: 437.66 ng/ml



#35 AR-1260-5

R.T.: 7.216 min
Delta R.T.: -0.007 min
Response: 79395459
Conc: 501.21 ng/ml