

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011124\
 Data File : PP062827.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jan 2024 08:07
 Operator : YP\AJ
 Sample : AR1242ICCC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 11 12:03:08 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 11 02:58:31 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.396	3.681	133.5E6	179.1E6	51.493	51.333
2)	SA Decachlor...	10.156	8.786	84480404	121.5E6	53.420	52.070
Target Compounds							
16)	L4 AR-1242-1	5.569	4.790	33470341	43732601	505.779	501.882
17)	L4 AR-1242-2	5.592	4.809	49529685	61253875	513.842	507.550
18)	L4 AR-1242-3	5.654	4.988	30782915	33930828	519.709	509.560
19)	L4 AR-1242-4	5.752	5.073	25413285	34746382	511.364	504.608
20)	L4 AR-1242-5	6.493	5.605	25496031	41477023	517.457	523.830



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011124\
Data File : PP062827.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jan 2024 08:07
Operator : YP\AJ
Sample : AR1242ICCC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 11 12:03:08 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 11 02:58:31 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.396 min
Delta R.T.: 0.000 min
Response: 133535793
Conc: 51.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICCC500




#1 Tetrachloro-m-xylene

R.T.: 3.681 min
Delta R.T.: 0.000 min
Response: 179109783
Conc: 51.33 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.156 min
Delta R.T.: -0.005 min
Response: 84480404
Conc: 53.42 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.786 min
Delta R.T.: -0.002 min
Response: 121498504
Conc: 52.07 ng/ml



#16 AR-1242-1


R.T.: 5.569 min
Delta R.T.: -0.001 min
Response: 33470341
Conc: 505.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICCC500




#16 AR-1242-1

R.T.: 4.790 min
Delta R.T.: -0.001 min
Response: 43732601
Conc: 501.88 ng/ml




#17 AR-1242-2

R.T.: 5.592 min
Delta R.T.: -0.002 min
Response: 49529685
Conc: 513.84 ng/ml



#17 AR-1242-2


R.T.: 4.809 min
Delta R.T.: -0.001 min
Response: 61253875
Conc: 507.55 ng/ml



#18 AR-1242-3


R.T.: 5.654 min
Delta R.T.: -0.002 min
Response: 30782915
Conc: 519.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICCC500




#18 AR-1242-3

R.T.: 4.988 min
Delta R.T.: -0.002 min
Response: 33930828
Conc: 509.56 ng/ml




#19 AR-1242-4

R.T.: 5.752 min
Delta R.T.: -0.001 min
Response: 25413285
Conc: 511.36 ng/ml



#19 AR-1242-4


R.T.: 5.073 min
Delta R.T.: -0.001 min
Response: 34746382
Conc: 504.61 ng/ml



#20 AR-1242-5

R.T.: 6.493 min
Delta R.T.: -0.001 min
Response: 25496031
Conc: 517.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICCC500



#20 AR-1242-5

R.T.: 5.605 min
Delta R.T.: -0.002 min
Response: 41477023
Conc: 523.83 ng/ml