

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011124\
 Data File : PP062842.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jan 2024 12:32
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 11 13:07:29 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 11 02:58:31 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.395	3.681	138.3E6	185.4E6	53.346	53.147
2)	SA Decachlor...	10.155	8.785	91353450	126.8E6	57.766	54.331
Target Compounds							
16)	L4 AR-1242-1	5.569	4.789	34598971	46342388	522.834	531.833
17)	L4 AR-1242-2	5.591	4.808	51386400	64792197	533.105	536.869
18)	L4 AR-1242-3	5.653	4.988	31849626	35774209	537.719	537.244
19)	L4 AR-1242-4	5.752	5.073	26140854	36347047	526.004	527.854
20)	L4 AR-1242-5	6.493	5.606	26303660	43604421	533.848	550.698



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011124\
Data File : PP062842.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jan 2024 12:32
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 11 13:07:29 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 11 02:58:31 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.395 min
Delta R.T.: 0.000 min
Response: 138343511
Conc: 53.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.681 min
Delta R.T.: -0.001 min
Response: 185438902
Conc: 53.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.155 min
Delta R.T.: -0.006 min
Response: 91353450
Conc: 57.77 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.785 min
Delta R.T.: -0.003 min
Response: 126776458
Conc: 54.33 ng/ml



#16 AR-1242-1


R.T.: 5.569 min
Delta R.T.: -0.002 min
Response: 34598971
Conc: 522.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.789 min
Delta R.T.: -0.002 min
Response: 46342388
Conc: 531.83 ng/ml




#17 AR-1242-2

R.T.: 5.591 min
Delta R.T.: -0.002 min
Response: 51386400
Conc: 533.10 ng/ml



#17 AR-1242-2


R.T.: 4.808 min
Delta R.T.: -0.002 min
Response: 64792197
Conc: 536.87 ng/ml



#18 AR-1242-3


R.T.: 5.653 min
Delta R.T.: -0.002 min
Response: 31849626
Conc: 537.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.988 min
Delta R.T.: -0.002 min
Response: 35774209
Conc: 537.24 ng/ml




#19 AR-1242-4

R.T.: 5.752 min
Delta R.T.: -0.002 min
Response: 26140854
Conc: 526.00 ng/ml



#19 AR-1242-4


R.T.: 5.073 min
Delta R.T.: -0.002 min
Response: 36347047
Conc: 527.85 ng/ml



#20 AR-1242-5

R.T.: 6.493 min
Delta R.T.: -0.002 min
Response: 26303660
Conc: 533.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.606 min
Delta R.T.: -0.002 min
Response: 43604421
Conc: 550.70 ng/ml