

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP011222.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jan 13 01:09:10 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP042823.D 1000 =PP042821.D 750 =PP042822.D
 250 =PP042824.D 50 =PP042825.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.396	2.412	2.358	1.944	2.530	2.328 E4	9.63
2) SA	Decachlorobiphenyl	1.406	1.390	1.363	1.122	1.738	1.404 E4	15.64
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	4.270	4.060	3.965	3.400	4.798	4.099 E2	12.36
22) L5	AR-1248-2	5.896	5.571	5.538	4.965	6.180	5.630 E2	8.08
23) L5	AR-1248-3	7.261	6.816	6.799	5.758	7.267	6.780 E2	9.08
24) L5	AR-1248-4	7.470	7.015	6.960	6.244	8.376	7.213 E2	10.88
25) L5	AR-1248-5	7.535	7.088	7.071	6.003	8.785	7.296 E2	13.77
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

500 =PP042823.D 1000 =PP042821.D 750 =PP042822.D
 250 =PP042824.D 50 =PP042825.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.640	2.605	2.539	2.173	2.999	2.591 E4	11.35
2) SA	Decachlorobiphenyl	2.003	1.864	1.879	1.739	2.358	1.969 E4	12.04
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP011222.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jan 13 01:09:10 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP042823.D 1000 =PP042821.D 750 =PP042822.D
 250 =PP042824.D 50 =PP042825.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	5.845	5.456	5.519	4.687	6.581	5.617 E2	12.21
22) L5	AR-1248-2	0.824	0.758	0.763	0.703	1.011	0.812 E3	14.71
23) L5	AR-1248-3	0.875	0.806	0.814	0.763	1.040	0.859 E3	12.62
24) L5	AR-1248-4	0.995	0.920	0.921	0.846	1.116	0.960 E3	10.64
25) L5	AR-1248-5	0.997	0.936	0.929	0.861	1.051	0.955 E3	7.56
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range