Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011222\
Data File : PP042809.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Jan 2022 13:18
Operator : AJ\MA
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 12 14:21:34 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 12 14:20:53 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1) SA	Tetrachlo...	4.875	4.044	2325322	2391925	95.141	95.458
2) SA	Decachlor...	10.805	9.381	1383227	1771723	95.753	90.420
	Target Compounds						
3) L1	AR-1016-1	6.178	5.309	681572	926051	894.814	904.481
4) L1	AR-1016-2	6.200	5.329	1067429	1324976	941.004	930.074
5) L1	AR-1016-3	6.267	5.523	677299	726537	963.463	909.032
6) L1	AR-1016-4	6.372	5.574	585635	555365	990.538	898.840
7) L1	AR-1016-5	6.687	5.806	547073	694288	1004.827	903.202
31) L7	AR-1260-1	7.859	6.905	829125	1152683	947.485	891.579
32) L7	AR-1260-2	8.123	7.102	909694	1361798	946.717	881.613
33) L7	AR-1260-3	8.489	7.260	677773	1270115	924.009	898.726
34) L7	AR-1260-4	8.718	7.745	797956	1022628	936.820	892.720
35) L7	AR-1260-5	9.035	7.992	1502723	2316685	945.950	924.734

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (Not Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011222\
Data File : PP042809.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Jan 2022 13:18
Operator : AJ\MA
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 12 14:21:34 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 12 14:20:53 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm







#1 Tetrachloro-m-xylene

R.T.: 4.875 min
Delta R.T.: 0.000 min
Response: 2325322
Conc: 95.14 ng/ml

Instrument :
ECD_P
ClientSampleId :



#1 Tetrachloro-m-xylene

R.T.: 4.044 min
Delta R.T.: 0.000 min
Response: 2391925
Conc: 95.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.805 min
Delta R.T.: 0.000 min
Response: 1383227
Conc: 95.75 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.381 min
Delta R.T.: 0.000 min
Response: 1771723
Conc: 90.42 ng/ml



#3 AR-1016-1

R.T.: 6.178 min
Delta R.T.: 0.000 min
Response: 681572
Conc: 894.81 ng/ml

Instrument :
ECD_P
ClientSampleId :



#3 AR-1016-1

R.T.: 5.309 min
Delta R.T.: 0.000 min
Response: 926051
Conc: 904.48 ng/ml



#4 AR-1016-2

R.T.: 6.200 min
Delta R.T.: 0.000 min
Response: 1067429
Conc: 941.00 ng/ml



#4 AR-1016-2

R.T.: 5.329 min
Delta R.T.: 0.000 min
Response: 1324976
Conc: 930.07 ng/ml



#5 AR-1016-3

R.T.: 6.267 min
Delta R.T.: 0.000 min
Response: 677299
Conc: 963.46 ng/ml

Instrument :
ECD_P
ClientSampleId :



#5 AR-1016-3

R.T.: 5.523 min
Delta R.T.: 0.000 min
Response: 726537
Conc: 909.03 ng/ml



#6 AR-1016-4

R.T.: 6.372 min
Delta R.T.: 0.001 min
Response: 585635
Conc: 990.54 ng/ml



#6 AR-1016-4

R.T.: 5.574 min
Delta R.T.: 0.000 min
Response: 555365
Conc: 898.84 ng/ml



#7 AR-1016-5

R.T.: 6.687 min
Delta R.T.: 0.000 min
Response: 547073
Conc: 1004.83 ng/m

Instrument :
ECD_P
ClientSampleId :



#7 AR-1016-5

R.T.: 5.806 min
Delta R.T.: 0.000 min
Response: 694288
Conc: 903.20 ng/ml



#31 AR-1260-1

R.T.: 7.859 min
Delta R.T.: 0.000 min
Response: 829125
Conc: 947.49 ng/ml



#31 AR-1260-1

R.T.: 6.905 min
Delta R.T.: 0.000 min
Response: 1152683
Conc: 891.58 ng/ml



#32 AR-1260-2

R.T.: 8.123 min
Delta R.T.: 0.000 min
Response: 909694
Conc: 946.72 ng/ml

Instrument :
ECD_P
ClientSampleId :



#32 AR-1260-2

R.T.: 7.102 min
Delta R.T.: 0.000 min
Response: 1361798
Conc: 881.61 ng/ml



#33 AR-1260-3

R.T.: 8.489 min
Delta R.T.: 0.000 min
Response: 677773
Conc: 924.01 ng/ml



#33 AR-1260-3

R.T.: 7.260 min
Delta R.T.: 0.000 min
Response: 1270115
Conc: 898.73 ng/ml



#34 AR-1260-4

R.T.: 8.718 min
Delta R.T.: 0.000 min
Response: 797956
Conc: 936.82 ng/ml

Instrument :
ECD_P
ClientSampleId :



#34 AR-1260-4

R.T.: 7.745 min
Delta R.T.: 0.000 min
Response: 1022628
Conc: 892.72 ng/ml



#35 AR-1260-5

R.T.: 9.035 min
Delta R.T.: 0.000 min
Response: 1502723
Conc: 945.95 ng/ml



#35 AR-1260-5

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 2316685
Conc: 924.73 ng/ml