

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011224\
 Data File : PP062872.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jan 2024 09:16
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 13 05:42:27 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 11 02:58:31 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.396	3.681	142.9E6	185.5E6	55.099	53.164
2) SA Decachlor...	10.152	8.781	94255628	131.8E6	59.601	56.497
Target Compounds						
3) L1 AR-1016-1	5.569	4.788	44456532	57668401	556.390	537.842
4) L1 AR-1016-2	5.591	4.807	65249933	81247183	549.878	543.328
5) L1 AR-1016-3	5.653	4.987	40387068	44165270	558.625	539.596
6) L1 AR-1016-4	5.752	5.029	33051267	31357278	538.945	465.795
7) L1 AR-1016-5	6.048	5.246	33554941	51428794	541.176	572.162
31) L7 AR-1260-1	7.180	6.295	59989444	90260925	580.708	535.105
32) L7 AR-1260-2	7.437	6.486	63979184	105.8E6	566.489	544.367
33) L7 AR-1260-3	7.799	6.642	44460561	98557446	522.881	532.607
34) L7 AR-1260-4	8.025	7.119	48121962	79855920	560.043	537.944
35) L7 AR-1260-5	8.344	7.363	90602232	179.4E6	570.313	566.308



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011224\
Data File : PP062872.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jan 2024 09:16
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 13 05:42:27 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 11 02:58:31 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.396 min
Delta R.T.: 0.000 min
Response: 142887404
Conc: 55.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.681 min
Delta R.T.: -0.001 min
Response: 185498043
Conc: 53.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.152 min
Delta R.T.: -0.009 min
Response: 94255628
Conc: 59.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.781 min
Delta R.T.: -0.007 min
Response: 131829844
Conc: 56.50 ng/ml



#3 AR-1016-1


R.T.: 5.569 min
Delta R.T.: -0.002 min
Response: 44456532
Conc: 556.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.788 min
Delta R.T.: -0.003 min
Response: 57668401
Conc: 537.84 ng/ml




#4 AR-1016-2

R.T.: 5.591 min
Delta R.T.: -0.002 min
Response: 65249933
Conc: 549.88 ng/ml



#4 AR-1016-2


R.T.: 4.807 min
Delta R.T.: -0.003 min
Response: 81247183
Conc: 543.33 ng/ml



#5 AR-1016-3


R.T.: 5.653 min
Delta R.T.: -0.002 min
Response: 40387068
Conc: 558.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.987 min
Delta R.T.: -0.003 min
Response: 44165270
Conc: 539.60 ng/ml




#6 AR-1016-4

R.T.: 5.752 min
Delta R.T.: -0.002 min
Response: 33051267
Conc: 538.94 ng/ml



#6 AR-1016-4


R.T.: 5.029 min
Delta R.T.: -0.002 min
Response: 31357278
Conc: 465.80 ng/ml



#7 AR-1016-5


R.T.: 6.048 min
Delta R.T.: -0.002 min
Response: 33554941
Conc: 541.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.246 min
Delta R.T.: -0.003 min
Response: 51428794
Conc: 572.16 ng/ml




#31 AR-1260-1

R.T.: 7.180 min
Delta R.T.: -0.003 min
Response: 59989444
Conc: 580.71 ng/ml



#31 AR-1260-1


R.T.: 6.295 min
Delta R.T.: -0.004 min
Response: 90260925
Conc: 535.10 ng/ml



#32 AR-1260-2


R.T.: 7.437 min
Delta R.T.: -0.004 min
Response: 63979184
Conc: 566.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.486 min
Delta R.T.: -0.005 min
Response: 105824611
Conc: 544.37 ng/ml




#33 AR-1260-3

R.T.: 7.799 min
Delta R.T.: -0.004 min
Response: 44460561
Conc: 522.88 ng/ml



#33 AR-1260-3


R.T.: 6.642 min
Delta R.T.: -0.004 min
Response: 98557446
Conc: 532.61 ng/ml



#34 AR-1260-4


R.T.: 8.025 min
Delta R.T.: -0.005 min
Response: 48121962
Conc: 560.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.119 min
Delta R.T.: -0.005 min
Response: 79855920
Conc: 537.94 ng/ml



#35 AR-1260-5

R.T.: 8.344 min
Delta R.T.: -0.005 min
Response: 90602232
Conc: 570.31 ng/ml



#35 AR-1260-5

R.T.: 7.363 min
Delta R.T.: -0.005 min
Response: 179436221
Conc: 566.31 ng/ml