

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011224\
 Data File : PP062873.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jan 2024 09:33
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 13 05:42:54 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 11 02:58:31 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.395	3.680	123.2E6	166.2E6	47.501	47.637
2) SA Decachlor...	10.152	8.782	85318051	120.1E6	53.950	51.464
Target Compounds						
16) L4 AR-1242-1	5.568	4.788	32255836	42072630	487.426	482.832
17) L4 AR-1242-2	5.590	4.807	47642384	58921600	494.263	488.225
18) L4 AR-1242-3	5.653	4.987	29461111	32504937	497.393	488.147
19) L4 AR-1242-4	5.751	5.072	24133670	33002328	485.615	479.280
20) L4 AR-1242-5	6.491	5.604	24517786	38966887	497.603	492.129



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011224\
Data File : PP062873.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jan 2024 09:33
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 13 05:42:54 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 11 02:58:31 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.395 min
Delta R.T.: -0.001 min
Response: 123183983
Conc: 47.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.680 min
Delta R.T.: -0.002 min
Response: 166214409
Conc: 47.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.152 min
Delta R.T.: -0.009 min
Response: 85318051
Conc: 53.95 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.782 min
Delta R.T.: -0.006 min
Response: 120084432
Conc: 51.46 ng/ml



#16 AR-1242-1


R.T.: 5.568 min
Delta R.T.: -0.003 min
Response: 32255836
Conc: 487.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.788 min
Delta R.T.: -0.003 min
Response: 42072630
Conc: 482.83 ng/ml




#17 AR-1242-2

R.T.: 5.590 min
Delta R.T.: -0.003 min
Response: 47642384
Conc: 494.26 ng/ml



#17 AR-1242-2


R.T.: 4.807 min
Delta R.T.: -0.003 min
Response: 58921600
Conc: 488.22 ng/ml



#18 AR-1242-3


R.T.: 5.653 min
Delta R.T.: -0.003 min
Response: 29461111
Conc: 497.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.987 min
Delta R.T.: -0.003 min
Response: 32504937
Conc: 488.15 ng/ml




#19 AR-1242-4

R.T.: 5.751 min
Delta R.T.: -0.003 min
Response: 24133670
Conc: 485.62 ng/ml



#19 AR-1242-4


R.T.: 5.072 min
Delta R.T.: -0.003 min
Response: 33002328
Conc: 479.28 ng/ml



#20 AR-1242-5

R.T.: 6.491 min
Delta R.T.: -0.004 min
Response: 24517786
Conc: 497.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.604 min
Delta R.T.: -0.003 min
Response: 38966887
Conc: 492.13 ng/ml