

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011422\
 Data File : PP042925.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Jan 2022 12:28
 Operator : AJ\MA
 Sample : PB142027BS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB142027BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 15 01:25:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 13 19:45:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.871	4.042	484494	522346	20.476	20.914
2) SA Decachlor...	10.804	9.378	341386	442770	23.353	22.698
Target Compounds						
3) L1 AR-1016-1	6.175	5.307	347008	480529	467.426	469.375
4) L1 AR-1016-2	6.198	5.327	509883	663143	443.401	464.337
5) L1 AR-1016-3	6.264	5.521	318466	370912	447.413	463.743
6) L1 AR-1016-4	6.369	5.572	269867	285153	445.060	458.808
7) L1 AR-1016-5	6.684	5.803	247328	354707	432.041	475.713
31) L7 AR-1260-1	7.858	6.904	411240	646432	431.267	500.112
32) L7 AR-1260-2	8.122	7.101	471056	754711	456.100	483.183
33) L7 AR-1260-3	8.488	7.258	298030	690307	395.008	477.629
34) L7 AR-1260-4	8.716	7.743	354081	490688	415.689	433.108
35) L7 AR-1260-5	9.033	7.991	647041	1065655	393.256	424.897



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011422\
Data File : PP042925.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Jan 2022 12:28
Operator : AJ\MA
Sample : PB142027BS
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB142027BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 15 01:25:38 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 13 19:45:07 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.871 min
Delta R.T.: -0.004 min
Response: 484494
Conc: 20.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142027BS




#1 Tetrachloro-m-xylene

R.T.: 4.042 min
Delta R.T.: -0.002 min
Response: 522346
Conc: 20.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.804 min
Delta R.T.: 0.000 min
Response: 341386
Conc: 23.35 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.378 min
Delta R.T.: -0.002 min
Response: 442770
Conc: 22.70 ng/ml



#3 AR-1016-1


R.T.: 6.175 min
Delta R.T.: -0.002 min
Response: 347008
Conc: 467.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142027BS




#3 AR-1016-1

R.T.: 5.307 min
Delta R.T.: -0.002 min
Response: 480529
Conc: 469.38 ng/ml




#4 AR-1016-2

R.T.: 6.198 min
Delta R.T.: -0.003 min
Response: 509883
Conc: 443.40 ng/ml



#4 AR-1016-2


R.T.: 5.327 min
Delta R.T.: -0.001 min
Response: 663143
Conc: 464.34 ng/ml



#5 AR-1016-3


R.T.: 6.264 min
Delta R.T.: -0.002 min
Response: 318466
Conc: 447.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142027BS




#5 AR-1016-3

R.T.: 5.521 min
Delta R.T.: -0.001 min
Response: 370912
Conc: 463.74 ng/ml




#6 AR-1016-4

R.T.: 6.369 min
Delta R.T.: -0.002 min
Response: 269867
Conc: 445.06 ng/ml



#6 AR-1016-4


R.T.: 5.572 min
Delta R.T.: -0.002 min
Response: 285153
Conc: 458.81 ng/ml



#7 AR-1016-5


R.T.: 6.684 min
Delta R.T.: -0.002 min
Response: 247328
Conc: 432.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142027BS




#7 AR-1016-5

R.T.: 5.803 min
Delta R.T.: -0.002 min
Response: 354707
Conc: 475.71 ng/ml




#31 AR-1260-1

R.T.: 7.858 min
Delta R.T.: -0.002 min
Response: 411240
Conc: 431.27 ng/ml



#31 AR-1260-1


R.T.: 6.904 min
Delta R.T.: -0.001 min
Response: 646432
Conc: 500.11 ng/ml



#32 AR-1260-2


R.T.: 8.122 min
Delta R.T.: 0.000 min
Response: 471056
Conc: 456.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142027BS




#32 AR-1260-2

R.T.: 7.101 min
Delta R.T.: -0.001 min
Response: 754711
Conc: 483.18 ng/ml




#33 AR-1260-3

R.T.: 8.488 min
Delta R.T.: -0.001 min
Response: 298030
Conc: 395.01 ng/ml



#33 AR-1260-3


R.T.: 7.258 min
Delta R.T.: -0.002 min
Response: 690307
Conc: 477.63 ng/ml



#34 AR-1260-4


R.T.: 8.716 min
Delta R.T.: -0.001 min
Response: 354081
Conc: 415.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142027BS




#34 AR-1260-4

R.T.: 7.743 min
Delta R.T.: -0.002 min
Response: 490688
Conc: 433.11 ng/ml



#35 AR-1260-5

R.T.: 9.033 min
Delta R.T.: -0.002 min
Response: 647041
Conc: 393.26 ng/ml



#35 AR-1260-5

R.T.: 7.991 min
Delta R.T.: 0.000 min
Response: 1065655
Conc: 424.90 ng/ml