

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011422\
 Data File : PP042950.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Jan 2022 21:20
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/17/2022
 Supervised By :Yogesh Patel 01/18/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 15 01:35:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 13 19:45:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.872	4.043	1125409	1271151	47.563	50.895
2) SA Decachlor...	10.802	9.378	787386	1017593	53.862m	52.167
Target Compounds						
26) L6 AR-1254-1	7.081	6.185	358257	733936	456.571	473.291
27) L6 AR-1254-2	7.310	6.344	561648	637170	446.252	469.352
28) L6 AR-1254-3	7.688	6.768	603203	1003403	464.932	469.113
29) L6 AR-1254-4	7.982	7.007	421816	614649	482.611	492.667
30) L6 AR-1254-5	8.410	7.438	445459	833532	475.453	488.826

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011422\
Data File : PP042950.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Jan 2022 21:20
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/17/2022
Supervised By :Yogesh Patel 01/18/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 15 01:35:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 13 19:45:07 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.872 min
Delta R.T.: -0.003 min
Response: 1125409
Conc: 47.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/17/2022
Supervised By :Yogesh Patel 01/18/2022




#1 Tetrachloro-m-xylene

R.T.: 4.043 min
Delta R.T.: -0.001 min
Response: 1271151
Conc: 50.89 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.802 min
Delta R.T.: -0.002 min
Response: 787386
Conc: 53.86 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.378 min
Delta R.T.: -0.002 min
Response: 1017593
Conc: 52.17 ng/ml




#26 AR-1254-1

R.T.: 7.081 min
Delta R.T.: 0.000 min
Response: 358257
Conc: 456.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/17/2022
Supervised By :Yogesh Patel 01/18/2022




#26 AR-1254-1

R.T.: 6.185 min
Delta R.T.: 0.000 min
Response: 733936
Conc: 473.29 ng/ml




#27 AR-1254-2

R.T.: 7.310 min
Delta R.T.: 0.000 min
Response: 561648
Conc: 446.25 ng/ml



#27 AR-1254-2

R.T.: 6.344 min
Delta R.T.: 0.000 min
Response: 637170
Conc: 469.35 ng/ml




#28 AR-1254-3

R.T.: 7.688 min
Delta R.T.: 0.000 min
Response: 603203
Conc: 464.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/17/2022
Supervised By :Yogesh Patel 01/18/2022




#28 AR-1254-3

R.T.: 6.768 min
Delta R.T.: 0.000 min
Response: 1003403
Conc: 469.11 ng/ml




#29 AR-1254-4

R.T.: 7.982 min
Delta R.T.: 0.000 min
Response: 421816
Conc: 482.61 ng/ml



#29 AR-1254-4

R.T.: 7.007 min
Delta R.T.: -0.001 min
Response: 614649
Conc: 492.67 ng/ml




#30 AR-1254-5

R.T.: 8.410 min
Delta R.T.: 0.000 min
Response: 445459
Conc: 475.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/17/2022
Supervised By :Yogesh Patel 01/18/2022



#30 AR-1254-5

R.T.: 7.438 min
Delta R.T.: -0.001 min
Response: 833532
Conc: 488.83 ng/ml