

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011519\
 Data File : PP028860.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Jan 2019 2:42
 Operator : EI\MA
 Sample : HSTDCCC750
 Misc : hexane
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 1/16/2019 1:35:23 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 16 07:11:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010319.M
 Quant Title : 8080.M
 QLast Update : Fri Jan 04 01:40:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.140	12.250	197.8E6	362.4E6	769.375	830.111
Target Compounds							
1) T	Dalapon	3.366	3.033	385.9E6	375.3E6	631.125	641.976
2) T	3,5-DICHL...	10.551	10.340	351.5E6	401.1E6	703.437	778.332
3) T	4-Nitroph...	11.718	11.396	304.7E6	170.6E6	770.402	867.523
5) T	DICAMBA	12.462	12.559	1122.4E6	1307.1E6	716.925	756.877m
6) T	MCPD	12.802	12.751	37836542	73150588	74.387m	79.089m
7) T	MCPA	13.046	13.141	56804667	114.8E6	71.333m	78.418
8) T	DICHLORPROP	13.658	13.675	255.9E6	332.4E6	791.119m	765.381
9) T	2,4-D	14.010	14.163	277.5E6	394.6E6	904.205m	808.980
10) T	Pentachlo...	14.431	14.778	5748.7E6	5014.5E6	732.199	771.225
11) T	2,4,5-TP ...	15.292	15.346	2080.5E6	1944.6E6	732.055	787.263
12) T	2,4,5-T	15.684	15.990	1906.1E6	1669.3E6	755.755	769.517
13) T	2,4-DB	16.447	16.588	161.4E6	225.6E6	872.325m	905.759m
14) T	DINOSEB	17.938	16.982	1528.7E6	1368.7E6	698.505	779.966
15) T	Picloram	17.705	18.367	2994.6E6	2723.3E6	701.801	733.235
16) T	DCPA	18.302	18.243	3058.7E6	2487.4E6	740.851	761.635

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011519\
Data File : PP028860.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Jan 2019 2:42
Operator : EI\MA
Sample : HSTDCCC750
Misc : hexane
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
1/16/2019 1:35:23 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 16 07:11:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010319.M
Quant Title : 8080.M
QLast Update : Fri Jan 04 01:40:06 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

