

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011621\
 Data File : PP032784.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Jan 2021 11:16
 Operator : DD\AJ
 Sample : M1114-06DL 40X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 S22-06-BDL

Manual Integrations
 APPROVED

Ankita
 1/18/2021 1:18:05 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 15 11:57:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 14 08:12:55 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds							
26)	L6	AR-1254-1	6.949	6.125	23962	31980	14.428m 15.620m
27)	L6	AR-1254-2	7.173	6.285	85195	67611	33.875 38.498
28)	L6	AR-1254-3	7.552	6.705	146572	143047	54.400 49.808
29)	L6	AR-1254-4	7.846	6.944	168465	129414	85.523 78.152
30)	L6	AR-1254-5	8.272	7.371	360551	346960	172.889 143.439
41)	L9	AR-1268-1	9.182	8.205	5711457	4643824	862.712 823.583
42)	L9	AR-1268-2	9.271	8.271	4288167	3832623	711.853 706.980
43)	L9	AR-1268-3	9.482	8.475	4714948	4034103	880.813 849.616
44)	L9	AR-1268-4	9.886	8.767	1281665	1042649	624.945 576.473
45)	L9	AR-1268-5	10.267	9.046	12603311	11749114	833.753 811.502

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011621\
Data File : PP032784.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Jan 2021 11:16
Operator : DD\AJ
Sample : M1114-06DL 40X
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampled :
S22-06-BDL


Manual Integrations
APPROVED

Ankita
1/18/2021 1:18:05 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 15 11:57:47 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011421.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 14 08:12:55 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 4.741 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
S22-06-BDL


Manual Integrations
APPROVED

Ankita
1/18/2021 1:18:05 PM




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 4.013 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 10.577 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 9.296 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 6.949 min
Delta R.T.: 0.003 min
Response: 23962
Conc: 14.43 ng/ml m

Instrument :
ECD_P
ClientSampleId :
S22-06-BDL


Manual Integrations
APPROVED

Ankita
1/18/2021 1:18:05 PM




#26 AR-1254-1

R.T.: 6.125 min
Delta R.T.: -0.006 min
Response: 31980
Conc: 15.62 ng/ml m




#27 AR-1254-2

R.T.: 7.173 min
Delta R.T.: -0.002 min
Response: 85195
Conc: 33.88 ng/ml



#27 AR-1254-2

R.T.: 6.285 min
Delta R.T.: -0.004 min
Response: 67611
Conc: 38.50 ng/ml




#28 AR-1254-3

R.T.: 7.552 min
Delta R.T.: -0.001 min
Response: 146572
Conc: 54.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
S22-06-BDL


Manual Integrations
APPROVED

Ankita
1/18/2021 1:18:05 PM




#28 AR-1254-3

R.T.: 6.705 min
Delta R.T.: -0.005 min
Response: 143047
Conc: 49.81 ng/ml




#29 AR-1254-4

R.T.: 7.846 min
Delta R.T.: -0.001 min
Response: 168465
Conc: 85.52 ng/ml



#29 AR-1254-4

R.T.: 6.944 min
Delta R.T.: -0.005 min
Response: 129414
Conc: 78.15 ng/ml




#30 AR-1254-5

R.T.: 8.272 min
Delta R.T.: -0.002 min
Response: 360551
Conc: 172.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
S22-06-BDL


Manual Integrations
APPROVED

Ankita
1/18/2021 1:18:05 PM




#30 AR-1254-5

R.T.: 7.371 min
Delta R.T.: -0.005 min
Response: 346960
Conc: 143.44 ng/ml




#41 AR-1268-1

R.T.: 9.182 min
Delta R.T.: -0.003 min
Response: 5711457
Conc: 862.71 ng/ml



#41 AR-1268-1

R.T.: 8.205 min
Delta R.T.: -0.005 min
Response: 4643824
Conc: 823.58 ng/ml




#42 AR-1268-2

R.T.: 9.271 min
Delta R.T.: -0.003 min
Response: 4288167
Conc: 711.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
S22-06-BDL


Manual Integrations
APPROVED

Ankita
1/18/2021 1:18:05 PM




#42 AR-1268-2

R.T.: 8.271 min
Delta R.T.: -0.006 min
Response: 3832623
Conc: 706.98 ng/ml




#43 AR-1268-3

R.T.: 9.482 min
Delta R.T.: -0.003 min
Response: 4714948
Conc: 880.81 ng/ml



#43 AR-1268-3

R.T.: 8.475 min
Delta R.T.: -0.006 min
Response: 4034103
Conc: 849.62 ng/ml




#44 AR-1268-4

R.T.: 9.886 min
Delta R.T.: -0.003 min
Response: 1281665
Conc: 624.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
S22-06-BDL


Manual Integrations
APPROVED

Ankita
1/18/2021 1:18:05 PM




#44 AR-1268-4

R.T.: 8.767 min
Delta R.T.: -0.006 min
Response: 1042649
Conc: 576.47 ng/ml



#45 AR-1268-5

R.T.: 10.267 min
Delta R.T.: -0.003 min
Response: 12603311
Conc: 833.75 ng/ml



#45 AR-1268-5

R.T.: 9.046 min
Delta R.T.: -0.006 min
Response: 11749114
Conc: 811.50 ng/ml