

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011621\
 Data File : PP032787.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Jan 2021 12:06
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 15 14:00:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 14 08:12:55 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.740	4.009	3315641	2589722	57.087	55.053
2) SA Decachlor...	10.572	9.286	2283912	2112358	57.889	56.386
Target Compounds						
3) L1 AR-1016-1	6.041	5.257	926715	740392	570.432	545.082
4) L1 AR-1016-2	6.064	5.278	1392119	1081560	563.384	533.234
5) L1 AR-1016-3	6.130	5.469	839620	601670	565.191	550.666
6) L1 AR-1016-4	6.235	5.520	710299	458186	570.685	565.994
7) L1 AR-1016-5	6.549	5.747	671978	440896	592.121	409.205 #
31) L7 AR-1260-1	7.719	6.839	1067256	1021098	559.839	535.276
32) L7 AR-1260-2	7.983	7.037	1375872	1254978	569.676	550.704
33) L7 AR-1260-3	8.348	7.191	1162856	1163291	576.926	531.979
34) L7 AR-1260-4	8.576	7.672	1181105	1000531	573.227	547.542
35) L7 AR-1260-5	8.888	7.921	2582797	2383508	570.163	554.089



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011621\
Data File : PP032787.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Jan 2021 12:06
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 15 14:00:38 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011421.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 14 08:12:55 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.740 min
Delta R.T.: -0.001 min
Response: 3315641
Conc: 57.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.009 min
Delta R.T.: -0.003 min
Response: 2589722
Conc: 55.05 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.572 min
Delta R.T.: -0.005 min
Response: 2283912
Conc: 57.89 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.286 min
Delta R.T.: -0.010 min
Response: 2112358
Conc: 56.39 ng/ml



#3 AR-1016-1


R.T.: 6.041 min
Delta R.T.: -0.002 min
Response: 926715
Conc: 570.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.257 min
Delta R.T.: -0.006 min
Response: 740392
Conc: 545.08 ng/ml




#4 AR-1016-2

R.T.: 6.064 min
Delta R.T.: -0.002 min
Response: 1392119
Conc: 563.38 ng/ml



#4 AR-1016-2


R.T.: 5.278 min
Delta R.T.: -0.006 min
Response: 1081560
Conc: 533.23 ng/ml



#5 AR-1016-3


R.T.: 6.130 min
Delta R.T.: -0.002 min
Response: 839620
Conc: 565.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.469 min
Delta R.T.: -0.006 min
Response: 601670
Conc: 550.67 ng/ml




#6 AR-1016-4

R.T.: 6.235 min
Delta R.T.: -0.003 min
Response: 710299
Conc: 570.69 ng/ml



#6 AR-1016-4


R.T.: 5.520 min
Delta R.T.: -0.006 min
Response: 458186
Conc: 565.99 ng/ml



#7 AR-1016-5


R.T.: 6.549 min
Delta R.T.: -0.003 min
Response: 671978
Conc: 592.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.747 min
Delta R.T.: -0.007 min
Response: 440896
Conc: 409.20 ng/ml




#31 AR-1260-1

R.T.: 7.719 min
Delta R.T.: -0.003 min
Response: 1067256
Conc: 559.84 ng/ml



#31 AR-1260-1


R.T.: 6.839 min
Delta R.T.: -0.007 min
Response: 1021098
Conc: 535.28 ng/ml



#32 AR-1260-2


R.T.: 7.983 min
Delta R.T.: -0.003 min
Response: 1375872
Conc: 569.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 7.037 min
Delta R.T.: -0.008 min
Response: 1254978
Conc: 550.70 ng/ml




#33 AR-1260-3

R.T.: 8.348 min
Delta R.T.: -0.003 min
Response: 1162856
Conc: 576.93 ng/ml



#33 AR-1260-3


R.T.: 7.191 min
Delta R.T.: -0.008 min
Response: 1163291
Conc: 531.98 ng/ml



#34 AR-1260-4


R.T.: 8.576 min
Delta R.T.: -0.003 min
Response: 1181105
Conc: 573.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.672 min
Delta R.T.: -0.008 min
Response: 1000531
Conc: 547.54 ng/ml



#35 AR-1260-5

R.T.: 8.888 min
Delta R.T.: -0.003 min
Response: 2582797
Conc: 570.16 ng/ml



#35 AR-1260-5

R.T.: 7.921 min
Delta R.T.: -0.008 min
Response: 2383508
Conc: 554.09 ng/ml