

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011625\
 Data File : PP069133.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jan 2025 12:45
 Operator : YP\AJ
 Sample : Q1091-02 10X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 BC260516-1-2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/17/2025
 Supervised By :Ankita Jodhani 01/17/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 16 13:07:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 07 11:05:55 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.667	3.980	2373141	1766788	1.706	1.810
2) SA Decachlor...	10.512	9.101	2047571	1687119	1.834m	1.404m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011625\
Data File : PP069133.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jan 2025 12:45
Operator : YP\AJ
Sample : Q1091-02 10X
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
BC260516-1-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/17/2025
Supervised By :Ankita Jodhani 01/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 16 13:07:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 07 11:05:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.667 min
Delta R.T.: -0.005 min
Response: 2373141
Conc: 1.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
BC260516-1-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/17/2025
Supervised By :Ankita Jodhani 01/17/2025




#1 Tetrachloro-m-xylene

R.T.: 3.980 min
Delta R.T.: -0.003 min
Response: 1766788
Conc: 1.81 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.512 min
Delta R.T.: -0.013 min
Response: 2047571
Conc: 1.83 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.101 min
Delta R.T.: -0.010 min
Response: 1687119
Conc: 1.40 ng/ml m