

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011625\
 Data File : PP069166.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jan 2025 00:54
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 17 05:36:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 07 11:05:55 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.671	3.980	70665463	46903843	50.791	48.049
2) SA Decachlor...	10.517	9.102	51174170	54420329	45.836	45.301
Target Compounds						
21) L5 AR-1248-1	5.831	5.080	14105558	10903608	479.330	474.087
22) L5 AR-1248-2	6.104	5.319	20145961	15470080	467.197	453.115
23) L5 AR-1248-3	6.308	5.362	22325905	15943820	473.232	451.246
24) L5 AR-1248-4	6.708	5.537	25498175	18623534	474.819	452.448
25) L5 AR-1248-5	6.748	5.933	24491639	17746732	470.837	458.735



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011625\
Data File : PP069166.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jan 2025 00:54
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 17 05:36:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 07 11:05:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.671 min
Delta R.T.: 0.000 min
Response: 70665463
Conc: 50.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.980 min
Delta R.T.: -0.003 min
Response: 46903843
Conc: 48.05 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.517 min
Delta R.T.: -0.007 min
Response: 51174170
Conc: 45.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.102 min
Delta R.T.: -0.009 min
Response: 54420329
Conc: 45.30 ng/ml



#21 AR-1248-1


R.T.: 5.831 min
Delta R.T.: 0.000 min
Response: 14105558
Conc: 479.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.080 min
Delta R.T.: -0.006 min
Response: 10903608
Conc: 474.09 ng/ml




#22 AR-1248-2

R.T.: 6.104 min
Delta R.T.: 0.000 min
Response: 20145961
Conc: 467.20 ng/ml



#22 AR-1248-2


R.T.: 5.319 min
Delta R.T.: -0.006 min
Response: 15470080
Conc: 453.12 ng/ml



#23 AR-1248-3


R.T.: 6.308 min
Delta R.T.: 0.000 min
Response: 22325905
Conc: 473.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.362 min
Delta R.T.: -0.006 min
Response: 15943820
Conc: 451.25 ng/ml




#24 AR-1248-4

R.T.: 6.708 min
Delta R.T.: 0.000 min
Response: 25498175
Conc: 474.82 ng/ml



#24 AR-1248-4


R.T.: 5.537 min
Delta R.T.: -0.007 min
Response: 18623534
Conc: 452.45 ng/ml



#25 AR-1248-5

R.T.: 6.748 min
Delta R.T.: 0.000 min
Response: 24491639
Conc: 470.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.933 min
Delta R.T.: -0.007 min
Response: 17746732
Conc: 458.74 ng/ml