

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011719\
 Data File : PP028882.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Jan 2019 10:14
 Operator : EI\MA
 Sample : HSTDICC500
 Misc : hexane
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 17 12:17:43 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011719.M
 Quant Title : 8080.M
 QLast Update : Thu Jan 17 12:16:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.159	12.270	138.1E6	240.6E6	550.056	541.214
Target Compounds							
1) T	Dalapon	3.376	3.042	310.4E6	298.8E6	493.667	493.458
2) T	3,5-DICHL...	10.570	10.359	265.6E6	283.2E6	507.123	505.125
3) T	4-Nitroph...	11.740	11.418	246.7E6	115.0E6	487.390	483.273
5) T	DICAMBA	12.481	12.579	772.2E6	849.7E6	509.373	505.591
6) T	MCPD	12.821	12.767	24503194	46198539	48.177	49.380
7) T	MCPA	13.064	13.156	39490771	77386245	49.387	49.596
8) T	DICHLORPROP	13.679	13.695	179.5E6	223.6E6	526.154	512.411
9) T	2,4-D	14.035	14.187	163.2E6	271.9E6	431.404	510.819
10) T	Pentachlo...	14.452	14.798	4010.2E6	3444.6E6	524.321	528.319
11) T	2,4,5-TP ...	15.315	15.367	1424.5E6	1300.6E6	514.548	514.977
12) T	2,4,5-T	15.708	16.015	1295.2E6	1096.5E6	511.590	509.471
13) T	2,4-DB	16.475	16.615	116.5E6	146.4E6	518.156	505.164
14) T	DINOSEB	17.960	17.002	1204.5E6	936.5E6	523.868	510.283
15) T	Picloram	17.736	18.394	2154.4E6	1856.4E6	503.977	510.182
16) T	DCPA	18.323	18.265	2178.1E6	1691.3E6	522.727	529.444

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011719\
Data File : PP028882.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Jan 2019 10:14
Operator : EI\MA
Sample : HSTDICC500
Misc : hexane
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
HSTDICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 17 12:17:43 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011719.M
Quant Title : 8080.M
QLast Update : Thu Jan 17 12:16:19 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

