

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP011720.M
 Title : 8080.M
 Last Update : Sat Jan 18 00:30:15 2020
 Response Via : Initial Calibration

Calibration Files

750 =PP030585.D 1000 =PP030586.D 1500 =PP030587.D
 500 =PP030584.D 200 =PP030583.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	5.812	5.624	5.425	5.853	5.743	5.691 E5	3.03
2) T	3,5-DICHLOROBENZOI...	6.201	5.766	5.499	6.358	6.424	6.049 E5	6.62
3) T	4-Nitrophenol	4.561	4.317	4.164	4.662	4.485	4.438 E5	4.47
4) S	2,4-DCAA	3.796	3.526	3.305	3.817	3.955	3.680 E5	7.08
5) T	DICAMBA	1.517	1.429	1.372	1.545	1.560	1.485 E6	5.44
6) T	MCPP	9.624	9.006	8.498	9.920	9.984	9.406 E5	6.79
7) T	MCPA	1.370	1.278	1.213	1.408	1.467	1.347 E6	7.54
8) T	DICHLORPROP	3.649	3.395	3.191	3.855	3.984	3.615 E5	8.99
9) T	2,4-D	4.093	3.750	3.638	4.064	4.318	3.973 E5	6.93
10) T	Pentachlorophenol	5.909	5.522	5.242	6.002	5.857	5.706 E6	5.55
11) T	2,4,5-TP (SILVEX)	2.028	1.910	1.832	2.078	2.068	1.983 E6	5.44
12) T	2,4,5-T	1.976	1.869	1.803	2.011	1.959	1.924 E6	4.44
13) T	2,4-DB	2.661	2.488	2.378	2.758	2.824	2.622 E5	7.09
14) T	DINOSEB	1.175	1.141	1.151	1.142	0.997	1.121 E6	6.29
15) T	Picloram	2.674	2.625	2.639	2.581	2.304	2.565 E6	5.83
16) T	DCPA	2.669	2.524	2.426	2.714	2.647	2.596 E6	4.56

Signal #2 Calibration Files

750 =PP030585.D 1000 =PP030586.D 1500 =PP030587.D
 500 =PP030584.D 200 =PP030583.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	2.593	2.509	2.472	2.556	2.433	2.513 E5	2.55
2) T	3,5-DICHLOROBENZOI...	2.583	2.460	2.367	2.605	2.560	2.515 E5	3.95
3) T	4-Nitrophenol	1.068	1.033	1.014	1.028	0.812	0.991 E5	10.30
4) S	2,4-DCAA	1.941	1.850	1.784	1.949	1.896	1.884 E5	3.64
5) T	DICAMBA	7.649	7.344	7.147	7.653	7.502	7.459 E5	2.89
6) T	MCPP	4.110	4.052	4.071	3.850	2.651	3.747 E5	16.57
7) T	MCPA	6.221	6.086	6.147	6.228	5.227	5.982 E5	7.12
8) T	DICHLORPROP	1.873	1.787	1.722	1.905	1.891	1.836 E5	4.26
9) T	2,4-D	2.126	2.035	1.978	2.094	2.002	2.047 E5	3.01
10) T	Pentachlorophenol	3.054	2.909	2.765	3.072	2.883	2.937 E6	4.35
11) T	2,4,5-TP (SILVEX)	1.097	1.059	1.027	1.100	1.038	1.064 E6	3.14
12) T	2,4,5-T	9.559	9.357	9.251	9.392	8.403	9.193 E5	4.95
13) T	2,4-DB	1.362	1.307	1.260	1.352	1.328	1.322 E5	3.06
14) T	DINOSEB	6.112	6.100	6.217	5.799	4.715	5.789 E5	10.71
15) T	Picloram	1.558	1.552	1.578	1.462	1.209	1.472 E6	10.44
16) T	DCPA	1.494	1.433	1.372	1.511	1.452	1.452 E6	3.76

(#) = Out of Range