

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011720\
 Data File : PP030588.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Jan 2020 20:03
 Operator : DD\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP011720

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 00:33:44 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011720.M
 Quant Title : 8080.M
 QLast Update : Sat Jan 18 00:30:15 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
1 T	Dalapon	682.500	654.553	4.1	94	0.00
2 T	3,5-DICHLOROBENZOIC ACID	697.500	679.053	2.6	95	0.00
3 T	4-Nitrophenol	682.500	671.984	1.5	96	0.00
4 S	2,4-DCAA	750.000	715.522	4.6	92	0.00
5 T	DICAMBA	705.000	684.125	3.0	95	0.00
6 T	MCP	70.500	69.099	2.0	96	0.00
7 T	MCPA	69.750	68.131	2.3	96	0.00
8 T	DICHLORPROP	705.000	683.483	3.1	96	0.00
9 T	2,4-D	705.000	696.037	1.3	96	0.00
10 T	Pentachlorophenol	712.500	698.597	2.0	95	0.00
11 T	2,4,5-TP (SILVEX)	712.500	692.758	2.8	95	0.00
12 T	2,4,5-T	712.500	700.710	1.7	96	0.00
13 T	2,4-DB	712.500	692.883	2.8	96	0.00
14 T	DINOSEB	705.000	722.186	-2.4	98	0.00
15 T	Picloram	712.500	714.233	-0.2	96	0.00
16 T	DCPA	720.000	705.707	2.0	95	0.00

Signal #2

1 T	Dalapon	682.500	676.475	0.9	96	0.00
2 T	3,5-DICHLOROBENZOIC ACID	697.500	687.291	1.5	96	0.00
3 T	4-Nitrophenol	682.500	710.107	-4.0	97	0.00
4 S	2,4-DCAA	750.000	735.270	2.0	95	0.00
5 T	DICAMBA	705.000	690.433	2.1	96	0.00
6 T	MCP	70.500	72.865	-3.4	94	0.00
7 T	MCPA	69.750	68.616	1.6	95	0.00
8 T	DICHLORPROP	705.000	685.384	2.8	95	0.00
9 T	2,4-D	705.000	687.758	2.4	94	0.00
10 T	Pentachlorophenol	712.500	707.549	0.7	95	0.00
11 T	2,4,5-TP (SILVEX)	712.500	699.688	1.8	95	0.00
12 T	2,4,5-T	712.500	711.538	0.1	96	0.00
13 T	2,4-DB	712.500	696.097	2.3	95	0.00
14 T	DINOSEB	705.000	734.012	-4.1	99	0.00
15 T	Picloram	712.500	729.774	-2.4	97	0.00
16 T	DCPA	720.000	706.962	1.8	95	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011720\
Data File : PP030588.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Jan 2020 20:03
Operator : DD\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP011720

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 00:33:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011720.M
Quant Title : 8080.M
QLast Update : Sat Jan 18 00:30:15 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0