

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011722\
 Data File : PP043004.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Jan 2022 21:27
 Operator : AJ\MA
 Sample : PB142064BS
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 02:59:35 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 13 19:45:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.878	4.044	490443	531246	20.728	21.270
2)	SA Decachlor...	10.816	9.382	372312	489538	25.469	25.096
Target Compounds							
3)	L1 AR-1016-1	6.182	5.309	318617	474572	429.184	463.556
4)	L1 AR-1016-2	6.205	5.329	479316	653580	416.820	457.641
5)	L1 AR-1016-3	6.271	5.523	306492	363905	430.592	454.984
6)	L1 AR-1016-4	6.376	5.574	245145	279844	404.288	450.267
7)	L1 AR-1016-5	6.691	5.805	225218	351378	393.419	471.248
8)	L2 AR-1221-1	5.114	4.302	45730	44312	171.781	132.705
9)	L2 AR-1221-2	5.218	4.403	43290	58217	214.208	239.277
10)	L2 AR-1221-3	5.304	4.493	192177	240440	292.704	306.711
11)	L3 AR-1232-1	5.304	4.493	192177	240440	336.940	369.281
12)	L3 AR-1232-2	5.890	5.329	257651	653580	949.921	1042.196
13)	L3 AR-1232-3	6.205	5.523	479316	363905	981.999	1100.455
14)	L3 AR-1232-4	6.376	5.618	245145	337601	1025.977	1143.737
15)	L3 AR-1232-5	6.477	5.805	181317	351378	1154.044	1139.314
16)	L4 AR-1242-1	6.182	5.309	318617	474572	537.780	592.190
17)	L4 AR-1242-2	6.205	5.329	479316	653580	537.245	577.217
18)	L4 AR-1242-3	6.271	5.523	306492	363905	552.385	570.736
19)	L4 AR-1242-4	6.376	5.618	245145	337601	538.802	549.608
20)	L4 AR-1242-5	7.156	6.186	29276	280303	65.266	389.544 #
21)	L5 AR-1248-1	6.182	5.309	318617	474572	777.370	844.832
22)	L5 AR-1248-2	6.477	5.574	181317	279844	322.061	344.788
23)	L5 AR-1248-3	6.691	5.618	225218	337601	332.181	392.793
24)	L5 AR-1248-4	7.115	5.805	28642	351378	39.709	366.122 #
25)	L5 AR-1248-5	7.156	6.230	29276	42703	40.124	44.730
26)	L6 AR-1254-1	7.089	6.186	167862	280303	213.927	180.758
27)	L6 AR-1254-2	7.317	6.345	179461	270309	142.589	199.114 #
28)	L6 AR-1254-3	7.696	6.786f	92605	466815	71.377	218.246 #
29)	L6 AR-1254-4	7.989	7.009	57755	49918	66.079	40.012 #
30)	L6 AR-1254-5	8.417	7.440	564166	853477	602.153	500.523
31)	L7 AR-1260-1	7.866	6.906	418462	644748	438.840	498.809
32)	L7 AR-1260-2	8.130	7.103	459025	749984	444.451	480.158
33)	L7 AR-1260-3	8.497	7.260	301234	700725	399.254	484.837
34)	L7 AR-1260-4	8.725	7.745	368182	498337	432.245	439.860
35)	L7 AR-1260-5	9.043	7.993	662659	1098011	402.748	437.797
36)	L8 AR-1262-1	8.497	7.440	301234	853477	285.271	942.802 #
37)	L8 AR-1262-2	9.043	7.745	662659	498337	382.823	333.142
38)	L8 AR-1262-3	9.364f	8.281	439580	224277	476.911	201.023 #
39)	L8 AR-1262-4	9.421	8.345	160268	798496	297.108	391.696 #
40)	L8 AR-1262-5	10.083	8.846	178955	250959	294.503	270.638
41)	L9 AR-1268-1	9.364f	8.281	439580	224277	213.221	72.766 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011722\
 Data File : PP043004.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Jan 2022 21:27
 Operator : AJ\MA
 Sample : PB142064BS
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 02:59:35 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 13 19:45:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.438	8.345	102874	798496	53.862	280.746 #
43) L9	AR-1268-3	9.666	8.556	24485	20006	14.535	8.395 #
44) L9	AR-1268-4	10.083	8.846	178955	250959	283.202	246.099
45) L9	AR-1268-5	10.486	9.132	60010	84835	11.260	11.942



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011722\
Data File : PP043004.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Jan 2022 21:27
Operator : AJ\MA
Sample : PB142064BS
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 02:59:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 13 19:45:07 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.878 min
Delta R.T.: 0.003 min
Response: 490443
Conc: 20.73 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.044 min
Delta R.T.: 0.000 min
Response: 531246
Conc: 21.27 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.816 min
Delta R.T.: 0.001 min
Response: 372312
Conc: 25.47 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.382 min
Delta R.T.: 0.002 min
Response: 489538
Conc: 25.10 ng/ml



#3 AR-1016-1


R.T.: 6.182 min
Delta R.T.: 0.005 min
Response: 318617
Conc: 429.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.309 min
Delta R.T.: 0.000 min
Response: 474572
Conc: 463.56 ng/ml




#4 AR-1016-2

R.T.: 6.205 min
Delta R.T.: 0.004 min
Response: 479316
Conc: 416.82 ng/ml



#4 AR-1016-2


R.T.: 5.329 min
Delta R.T.: 0.000 min
Response: 653580
Conc: 457.64 ng/ml



#5 AR-1016-3


R.T.: 6.271 min
Delta R.T.: 0.005 min
Response: 306492
Conc: 430.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.523 min
Delta R.T.: 0.000 min
Response: 363905
Conc: 454.98 ng/ml




#6 AR-1016-4

R.T.: 6.376 min
Delta R.T.: 0.005 min
Response: 245145
Conc: 404.29 ng/ml



#6 AR-1016-4


R.T.: 5.574 min
Delta R.T.: 0.000 min
Response: 279844
Conc: 450.27 ng/ml



#7 AR-1016-5


R.T.: 6.691 min
Delta R.T.: 0.005 min
Response: 225218
Conc: 393.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.805 min
Delta R.T.: 0.000 min
Response: 351378
Conc: 471.25 ng/ml




#8 AR-1221-1

R.T.: 5.114 min
Delta R.T.: -0.003 min
Response: 45730
Conc: 171.78 ng/ml



#8 AR-1221-1


R.T.: 4.302 min
Delta R.T.: -0.001 min
Response: 44312
Conc: 132.71 ng/ml



#9 AR-1221-2


R.T.: 5.218 min
Delta R.T.: 0.005 min
Response: 43290
Conc: 214.21 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.403 min
Delta R.T.: 0.000 min
Response: 58217
Conc: 239.28 ng/ml




#10 AR-1221-3

R.T.: 5.304 min
Delta R.T.: 0.004 min
Response: 192177
Conc: 292.70 ng/ml



#10 AR-1221-3


R.T.: 4.493 min
Delta R.T.: 0.000 min
Response: 240440
Conc: 306.71 ng/ml



#11 AR-1232-1


R.T.: 5.304 min
Delta R.T.: 0.004 min
Response: 192177
Conc: 336.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.493 min
Delta R.T.: 0.000 min
Response: 240440
Conc: 369.28 ng/ml




#12 AR-1232-2

R.T.: 5.890 min
Delta R.T.: 0.004 min
Response: 257651
Conc: 949.92 ng/ml



#12 AR-1232-2


R.T.: 5.329 min
Delta R.T.: 0.000 min
Response: 653580
Conc: 1042.20 ng/ml



#13 AR-1232-3


R.T.: 6.205 min
Delta R.T.: 0.005 min
Response: 479316
Conc: 982.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.523 min
Delta R.T.: 0.000 min
Response: 363905
Conc: 1100.46 ng/ml




#14 AR-1232-4

R.T.: 6.376 min
Delta R.T.: 0.004 min
Response: 245145
Conc: 1025.98 ng/ml



#14 AR-1232-4


R.T.: 5.618 min
Delta R.T.: 0.000 min
Response: 337601
Conc: 1143.74 ng/ml



#15 AR-1232-5


R.T.: 6.477 min
Delta R.T.: 0.005 min
Response: 181317
Conc: 1154.04 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.805 min
Delta R.T.: 0.000 min
Response: 351378
Conc: 1139.31 ng/ml




#16 AR-1242-1

R.T.: 6.182 min
Delta R.T.: 0.006 min
Response: 318617
Conc: 537.78 ng/ml



#16 AR-1242-1


R.T.: 5.309 min
Delta R.T.: 0.000 min
Response: 474572
Conc: 592.19 ng/ml



#17 AR-1242-2


R.T.: 6.205 min
Delta R.T.: 0.006 min
Response: 479316
Conc: 537.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.329 min
Delta R.T.: 0.000 min
Response: 653580
Conc: 577.22 ng/ml




#18 AR-1242-3

R.T.: 6.271 min
Delta R.T.: 0.005 min
Response: 306492
Conc: 552.38 ng/ml



#18 AR-1242-3


R.T.: 5.523 min
Delta R.T.: 0.000 min
Response: 363905
Conc: 570.74 ng/ml



#19 AR-1242-4


R.T.: 6.376 min
Delta R.T.: 0.005 min
Response: 245145
Conc: 538.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.618 min
Delta R.T.: 0.000 min
Response: 337601
Conc: 549.61 ng/ml




#20 AR-1242-5

R.T.: 7.156 min
Delta R.T.: 0.007 min
Response: 29276
Conc: 65.27 ng/ml



#20 AR-1242-5


R.T.: 6.186 min
Delta R.T.: 0.000 min
Response: 280303
Conc: 389.54 ng/ml



#21 AR-1248-1


R.T.: 6.182 min
Delta R.T.: 0.006 min
Response: 318617
Conc: 777.37 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.309 min
Delta R.T.: 0.001 min
Response: 474572
Conc: 844.83 ng/ml




#22 AR-1248-2

R.T.: 6.477 min
Delta R.T.: 0.006 min
Response: 181317
Conc: 322.06 ng/ml



#22 AR-1248-2


R.T.: 5.574 min
Delta R.T.: 0.000 min
Response: 279844
Conc: 344.79 ng/ml



#23 AR-1248-3


R.T.: 6.691 min
Delta R.T.: 0.006 min
Response: 225218
Conc: 332.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.618 min
Delta R.T.: 0.000 min
Response: 337601
Conc: 392.79 ng/ml




#24 AR-1248-4

R.T.: 7.115 min
Delta R.T.: 0.006 min
Response: 28642
Conc: 39.71 ng/ml



#24 AR-1248-4


R.T.: 5.805 min
Delta R.T.: 0.000 min
Response: 351378
Conc: 366.12 ng/ml



#25 AR-1248-5


R.T.: 7.156 min
Delta R.T.: 0.007 min
Response: 29276
Conc: 40.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.230 min
Delta R.T.: 0.001 min
Response: 42703
Conc: 44.73 ng/ml




#26 AR-1254-1

R.T.: 7.089 min
Delta R.T.: 0.007 min
Response: 167862
Conc: 213.93 ng/ml



#26 AR-1254-1


R.T.: 6.186 min
Delta R.T.: 0.000 min
Response: 280303
Conc: 180.76 ng/ml



#27 AR-1254-2


R.T.: 7.317 min
Delta R.T.: 0.007 min
Response: 179461
Conc: 142.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.345 min
Delta R.T.: 0.000 min
Response: 270309
Conc: 199.11 ng/ml




#28 AR-1254-3

R.T.: 7.696 min
Delta R.T.: 0.007 min
Response: 92605
Conc: 71.38 ng/ml



#28 AR-1254-3


R.T.: 6.786 min
Delta R.T.: 0.017 min
Response: 466815
Conc: 218.25 ng/ml



#29 AR-1254-4


R.T.: 7.989 min
Delta R.T.: 0.007 min
Response: 57755
Conc: 66.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 7.009 min
Delta R.T.: 0.000 min
Response: 49918
Conc: 40.01 ng/ml




#30 AR-1254-5

R.T.: 8.417 min
Delta R.T.: 0.007 min
Response: 564166
Conc: 602.15 ng/ml



#30 AR-1254-5


R.T.: 7.440 min
Delta R.T.: 0.000 min
Response: 853477
Conc: 500.52 ng/ml



#31 AR-1260-1


R.T.: 7.866 min
Delta R.T.: 0.006 min
Response: 418462
Conc: 438.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.906 min
Delta R.T.: 0.000 min
Response: 644748
Conc: 498.81 ng/ml




#32 AR-1260-2

R.T.: 8.130 min
Delta R.T.: 0.007 min
Response: 459025
Conc: 444.45 ng/ml



#32 AR-1260-2


R.T.: 7.103 min
Delta R.T.: 0.000 min
Response: 749984
Conc: 480.16 ng/ml



#33 AR-1260-3


R.T.: 8.497 min
Delta R.T.: 0.007 min
Response: 301234
Conc: 399.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.260 min
Delta R.T.: 0.000 min
Response: 700725
Conc: 484.84 ng/ml




#34 AR-1260-4

R.T.: 8.725 min
Delta R.T.: 0.007 min
Response: 368182
Conc: 432.24 ng/ml



#34 AR-1260-4


R.T.: 7.745 min
Delta R.T.: 0.000 min
Response: 498337
Conc: 439.86 ng/ml



#35 AR-1260-5


R.T.: 9.043 min
Delta R.T.: 0.008 min
Response: 662659
Conc: 402.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.993 min
Delta R.T.: 0.001 min
Response: 1098011
Conc: 437.80 ng/ml




#36 AR-1262-1

R.T.: 8.497 min
Delta R.T.: 0.007 min
Response: 301234
Conc: 285.27 ng/ml



#36 AR-1262-1


R.T.: 7.440 min
Delta R.T.: 0.002 min
Response: 853477
Conc: 942.80 ng/ml



#37 AR-1262-2


R.T.: 9.043 min
Delta R.T.: 0.008 min
Response: 662659
Conc: 382.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.745 min
Delta R.T.: 0.001 min
Response: 498337
Conc: 333.14 ng/ml




#38 AR-1262-3

R.T.: 9.364 min
Delta R.T.: 0.020 min
Response: 439580
Conc: 476.91 ng/ml



#38 AR-1262-3


R.T.: 8.281 min
Delta R.T.: 0.002 min
Response: 224277
Conc: 201.02 ng/ml



#39 AR-1262-4


R.T.: 9.421 min
Delta R.T.: -0.008 min
Response: 160268
Conc: 297.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.345 min
Delta R.T.: 0.000 min
Response: 798496
Conc: 391.70 ng/ml




#40 AR-1262-5

R.T.: 10.083 min
Delta R.T.: 0.010 min
Response: 178955
Conc: 294.50 ng/ml



#40 AR-1262-5


R.T.: 8.846 min
Delta R.T.: 0.002 min
Response: 250959
Conc: 270.64 ng/ml



#41 AR-1268-1


R.T.: 9.364 min
Delta R.T.: 0.023 min
Response: 439580
Conc: 213.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.281 min
Delta R.T.: 0.001 min
Response: 224277
Conc: 72.77 ng/ml




#42 AR-1268-2

R.T.: 9.438 min
Delta R.T.: 0.004 min
Response: 102874
Conc: 53.86 ng/ml



#42 AR-1268-2


R.T.: 8.345 min
Delta R.T.: 0.000 min
Response: 798496
Conc: 280.75 ng/ml



#43 AR-1268-3


R.T.: 9.666 min
Delta R.T.: 0.009 min
Response: 24485
Conc: 14.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.556 min
Delta R.T.: 0.002 min
Response: 20006
Conc: 8.39 ng/ml




#44 AR-1268-4

R.T.: 10.083 min
Delta R.T.: 0.010 min
Response: 178955
Conc: 283.20 ng/ml



#44 AR-1268-4


R.T.: 8.846 min
Delta R.T.: 0.002 min
Response: 250959
Conc: 246.10 ng/ml



#45 AR-1268-5

R.T.: 10.486 min
Delta R.T.: 0.010 min
Response: 60010
Conc: 11.26 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.132 min
Delta R.T.: 0.001 min
Response: 84835
Conc: 11.94 ng/ml