

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011723\
 Data File : PP054641.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jan 2023 00:27
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 01:52:50 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 18 01:52:34 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.331	3.578	91389859	80316651	50.000	50.000
2) SA Decachlor...	10.095	8.602	67089275	69543034	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.364	5.470	35093891	41840874	500.000	500.000
27) L6 AR-1254-2	6.583	5.619	46945624	36146833	500.000	500.000
28) L6 AR-1254-3	6.952	6.024	43564672	57139440	500.000	500.000
29) L6 AR-1254-4	7.239	6.254	31471652	31598205	500.000	500.000
30) L6 AR-1254-5	7.660	6.673	34640510	47471662	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011723\
Data File : PP054641.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jan 2023 00:27
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 01:52:50 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011723.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 18 01:52:34 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.331 min
Delta R.T.: 0.000 min
Response: 91389859
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.578 min
Delta R.T.: 0.000 min
Response: 80316651
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.095 min
Delta R.T.: 0.000 min
Response: 67089275
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.602 min
Delta R.T.: 0.000 min
Response: 69543034
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 6.364 min
Delta R.T.: 0.000 min
Response: 35093891
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#26 AR-1254-1

R.T.: 5.470 min
Delta R.T.: 0.000 min
Response: 41840874
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 6.583 min
Delta R.T.: 0.000 min
Response: 46945624
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 5.619 min
Delta R.T.: 0.000 min
Response: 36146833
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 6.952 min
Delta R.T.: 0.000 min
Response: 43564672
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#28 AR-1254-3

R.T.: 6.024 min
Delta R.T.: 0.000 min
Response: 57139440
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 7.239 min
Delta R.T.: 0.000 min
Response: 31471652
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 6.254 min
Delta R.T.: 0.000 min
Response: 31598205
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 7.660 min
Delta R.T.: 0.000 min
Response: 34640510
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500



#30 AR-1254-5

R.T.: 6.673 min
Delta R.T.: 0.000 min
Response: 47471662
Conc: 500.00 ng/ml