

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011723\
 Data File : PP054653.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jan 2023 04:01
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP011723AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 04:13:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 18 04:10:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.331	3.578	90663027	79369913	50.022	50.337
2) SA Decachlor...	10.095	8.601	66534652	68801037	49.681	49.057
Target Compounds						
26) L6 AR-1254-1	6.364	5.470	36136237	41559752	511.269	488.817
27) L6 AR-1254-2	6.583	5.619	48745612	35783741	509.285	491.752
28) L6 AR-1254-3	6.951	6.023	43249086	55973337	498.482	494.571
29) L6 AR-1254-4	7.238	6.253	31134198	31031448	496.963	498.570
30) L6 AR-1254-5	7.660	6.673	34278068	47101197	489.834	497.009



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011723\
Data File : PP054653.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jan 2023 04:01
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP011723AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 04:13:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011723.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 18 04:10:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.331 min
Delta R.T.: 0.000 min
Response: 90663027
Conc: 50.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP011723AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.578 min
Delta R.T.: 0.000 min
Response: 79369913
Conc: 50.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.095 min
Delta R.T.: 0.000 min
Response: 66534652
Conc: 49.68 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.601 min
Delta R.T.: -0.002 min
Response: 68801037
Conc: 49.06 ng/ml



#26 AR-1254-1


R.T.: 6.364 min
Delta R.T.: 0.000 min
Response: 36136237
Conc: 511.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP011723AR1254




#26 AR-1254-1

R.T.: 5.470 min
Delta R.T.: 0.000 min
Response: 41559752
Conc: 488.82 ng/ml




#27 AR-1254-2

R.T.: 6.583 min
Delta R.T.: -0.001 min
Response: 48745612
Conc: 509.29 ng/ml



#27 AR-1254-2


R.T.: 5.619 min
Delta R.T.: 0.000 min
Response: 35783741
Conc: 491.75 ng/ml



#28 AR-1254-3


R.T.: 6.951 min
Delta R.T.: 0.000 min
Response: 43249086
Conc: 498.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP011723AR1254




#28 AR-1254-3

R.T.: 6.023 min
Delta R.T.: -0.001 min
Response: 55973337
Conc: 494.57 ng/ml




#29 AR-1254-4

R.T.: 7.238 min
Delta R.T.: 0.000 min
Response: 31134198
Conc: 496.96 ng/ml



#29 AR-1254-4


R.T.: 6.253 min
Delta R.T.: 0.000 min
Response: 31031448
Conc: 498.57 ng/ml



#30 AR-1254-5

R.T.: 7.660 min
Delta R.T.: 0.000 min
Response: 34278068
Conc: 489.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP011723AR1254



#30 AR-1254-5

R.T.: 6.673 min
Delta R.T.: 0.000 min
Response: 47101197
Conc: 497.01 ng/ml