

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011724\
 Data File : PP062957.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jan 2024 09:19
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 17 21:31:20 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 11 02:58:31 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.395	3.680	124.6E6	167.6E6	48.044	48.047
2) SA Decachlor...	10.155	8.783	88165317	122.1E6	55.750	52.337
Target Compounds						
21) L5 AR-1248-1	5.568	4.787	24798314	32165841	497.574	489.561
22) L5 AR-1248-2	5.843	5.029	37364097	45831684	474.218	470.217
23) L5 AR-1248-3	6.048	5.071	40822691	48319144	480.944	473.619
24) L5 AR-1248-4	6.453	5.246	41546579	56919217	487.313	476.034
25) L5 AR-1248-5	6.492	5.646	40902790	53831257	487.011	496.477

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011724\
Data File : PP062957.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jan 2024 09:19
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 17 21:31:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 11 02:58:31 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

