

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011725\
 Data File : PP069173.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jan 2025 08:48
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/21/2025
 Supervised By :Ankita Jodhani 01/21/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 17 10:11:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 07 11:05:55 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.668	3.980	72573892	48768172	52.163	49.959
2) SA Decachlor...	10.510	9.101	53684325	56145190	48.084	46.737
Target Compounds						
26) L6 AR-1254-1	6.680	5.891	29519809	28304026	554.339m	470.700
27) L6 AR-1254-2	6.898	6.039	39756591	24864821	488.973	464.773
28) L6 AR-1254-3	7.263	6.447	40643561	38054332	504.401	462.269
29) L6 AR-1254-4	7.545	6.675	31047649	22703960	509.820	504.531
30) L6 AR-1254-5	7.962	7.096	33499442	34759174	524.137m	465.874

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011725\
Data File : PP069173.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jan 2025 08:48
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/21/2025
Supervised By :Ankita Jodhani 01/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 17 10:11:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 07 11:05:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.668 min
Delta R.T.: -0.004 min
Response: 72573892
Conc: 52.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/21/2025
Supervised By :Ankita Jodhani 01/21/2025




#1 Tetrachloro-m-xylene

R.T.: 3.980 min
Delta R.T.: -0.003 min
Response: 48768172
Conc: 49.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.510 min
Delta R.T.: -0.014 min
Response: 53684325
Conc: 48.08 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.101 min
Delta R.T.: -0.010 min
Response: 56145190
Conc: 46.74 ng/ml




#26 AR-1254-1

R.T.: 6.680 min
Delta R.T.: -0.006 min
Response: 29519809
Conc: 554.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/21/2025
Supervised By :Ankita Jodhani 01/21/2025




#26 AR-1254-1

R.T.: 5.891 min
Delta R.T.: -0.007 min
Response: 28304026
Conc: 470.70 ng/ml




#27 AR-1254-2

R.T.: 6.898 min
Delta R.T.: -0.006 min
Response: 39756591
Conc: 488.97 ng/ml



#27 AR-1254-2

R.T.: 6.039 min
Delta R.T.: -0.006 min
Response: 24864821
Conc: 464.77 ng/ml




#28 AR-1254-3

R.T.: 7.263 min
Delta R.T.: -0.006 min
Response: 40643561
Conc: 504.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/21/2025
Supervised By :Ankita Jodhani 01/21/2025




#28 AR-1254-3

R.T.: 6.447 min
Delta R.T.: -0.006 min
Response: 38054332
Conc: 462.27 ng/ml




#29 AR-1254-4

R.T.: 7.545 min
Delta R.T.: -0.007 min
Response: 31047649
Conc: 509.82 ng/ml



#29 AR-1254-4

R.T.: 6.675 min
Delta R.T.: -0.007 min
Response: 22703960
Conc: 504.53 ng/ml




#30 AR-1254-5

R.T.: 7.962 min
Delta R.T.: -0.007 min
Response: 33499442
Conc: 524.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/21/2025
Supervised By :Ankita Jodhani 01/21/2025



#30 AR-1254-5

R.T.: 7.096 min
Delta R.T.: -0.007 min
Response: 34759174
Conc: 465.87 ng/ml