

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011725\
 Data File : PP069175.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jan 2025 09:20
 Operator : YP\AJ
 Sample : Q1094-03RE 10X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 3181

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/21/2025
 Supervised By :Ankita Jodhani 01/21/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 17 14:52:48 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 07 11:05:55 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.703f	3.977	803959	354682	0.578	0.363 #
2) SA Decachlor...	10.507	9.108	92347	402448	0.083m	0.335 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011725\
Data File : PP069175.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jan 2025 09:20
Operator : YP\AJ
Sample : Q1094-03RE 10X
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
3181


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/21/2025
Supervised By :Ankita Jodhani 01/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 17 14:52:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 07 11:05:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.703 min
Delta R.T.: 0.031 min
Response: 803959
Conc: 0.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
3181


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/21/2025
Supervised By :Ankita Jodhani 01/21/2025




#1 Tetrachloro-m-xylene

R.T.: 3.977 min
Delta R.T.: -0.006 min
Response: 354682
Conc: 0.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.507 min
Delta R.T.: -0.017 min
Response: 92347
Conc: 0.08 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.108 min
Delta R.T.: -0.004 min
Response: 402448
Conc: 0.34 ng/ml