

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011822\
 Data File : PP043050.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Jan 2022 11:32
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 19 01:19:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 13 19:45:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.875	4.041	1153843	1247833	48.765	49.961
2) SA Decachlor...	10.804	9.374	836430	1005153	57.217	51.529
Target Compounds						
3) L1 AR-1016-1	6.178	5.305	356098	527224	479.672	514.986
4) L1 AR-1016-2	6.200	5.325	542800	723754	472.027	506.777
5) L1 AR-1016-3	6.267	5.519	340398	406657	478.226	508.436
6) L1 AR-1016-4	6.372	5.570	274565	318888	452.807	513.087
7) L1 AR-1016-5	6.688	5.802	260723	368302	455.441	493.946
31) L7 AR-1260-1	7.860	6.901	425835	674634	446.573	521.930
32) L7 AR-1260-2	8.124	7.098	476941	776301	461.799	497.006
33) L7 AR-1260-3	8.491	7.255	368455	698672	488.348	483.417
34) L7 AR-1260-4	8.719	7.740	431743	577103	506.865	509.384
35) L7 AR-1260-5	9.036	7.988	789328	1251045	479.735	498.815



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011822\
Data File : PP043050.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Jan 2022 11:32
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 19 01:19:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 13 19:45:07 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.875 min
Delta R.T.: 0.000 min
Response: 1153843
Conc: 48.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.041 min
Delta R.T.: -0.003 min
Response: 1247833
Conc: 49.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.804 min
Delta R.T.: -0.011 min
Response: 836430
Conc: 57.22 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.374 min
Delta R.T.: -0.006 min
Response: 1005153
Conc: 51.53 ng/ml



#3 AR-1016-1


R.T.: 6.178 min
Delta R.T.: 0.001 min
Response: 356098
Conc: 479.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.305 min
Delta R.T.: -0.003 min
Response: 527224
Conc: 514.99 ng/ml




#4 AR-1016-2

R.T.: 6.200 min
Delta R.T.: 0.000 min
Response: 542800
Conc: 472.03 ng/ml



#4 AR-1016-2


R.T.: 5.325 min
Delta R.T.: -0.004 min
Response: 723754
Conc: 506.78 ng/ml



#5 AR-1016-3


R.T.: 6.267 min
Delta R.T.: 0.001 min
Response: 340398
Conc: 478.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.519 min
Delta R.T.: -0.003 min
Response: 406657
Conc: 508.44 ng/ml




#6 AR-1016-4

R.T.: 6.372 min
Delta R.T.: 0.001 min
Response: 274565
Conc: 452.81 ng/ml



#6 AR-1016-4


R.T.: 5.570 min
Delta R.T.: -0.004 min
Response: 318888
Conc: 513.09 ng/ml



#7 AR-1016-5


R.T.: 6.688 min
Delta R.T.: 0.001 min
Response: 260723
Conc: 455.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.802 min
Delta R.T.: -0.003 min
Response: 368302
Conc: 493.95 ng/ml




#31 AR-1260-1

R.T.: 7.860 min
Delta R.T.: 0.000 min
Response: 425835
Conc: 446.57 ng/ml



#31 AR-1260-1


R.T.: 6.901 min
Delta R.T.: -0.004 min
Response: 674634
Conc: 521.93 ng/ml



#32 AR-1260-2


R.T.: 8.124 min
Delta R.T.: 0.002 min
Response: 476941
Conc: 461.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 7.098 min
Delta R.T.: -0.004 min
Response: 776301
Conc: 497.01 ng/ml




#33 AR-1260-3

R.T.: 8.491 min
Delta R.T.: 0.001 min
Response: 368455
Conc: 488.35 ng/ml



#33 AR-1260-3


R.T.: 7.255 min
Delta R.T.: -0.005 min
Response: 698672
Conc: 483.42 ng/ml



#34 AR-1260-4


R.T.: 8.719 min
Delta R.T.: 0.001 min
Response: 431743
Conc: 506.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.740 min
Delta R.T.: -0.005 min
Response: 577103
Conc: 509.38 ng/ml



#35 AR-1260-5

R.T.: 9.036 min
Delta R.T.: 0.000 min
Response: 789328
Conc: 479.74 ng/ml



#35 AR-1260-5

R.T.: 7.988 min
Delta R.T.: -0.005 min
Response: 1251045
Conc: 498.82 ng/ml