

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011822\
 Data File : PP043066.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Jan 2022 16:34
 Operator : AJ\MA
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 19 04:39:34 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 13 19:45:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.873	4.039	1181186	1261238	49.921	50.498
2) SA Decachlor...	10.800	9.371	831144	1023892	56.856	52.489
Target Compounds						
16) L4 AR-1242-1	6.176	5.303	277594	410133	468.539	511.781
17) L4 AR-1242-2	6.198	5.323	418585	565226	469.174	499.186
18) L4 AR-1242-3	6.265	5.517	263568	315159	475.023	494.284
19) L4 AR-1242-4	6.369	5.612	211169	295847	464.126	481.634
20) L4 AR-1242-5	7.148	6.179	208747	369655	465.371	513.720



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011822\
Data File : PP043066.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Jan 2022 16:34
Operator : AJ\MA
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 19 04:39:34 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 13 19:45:07 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.873 min
Delta R.T.: -0.002 min
Response: 1181186
Conc: 49.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.039 min
Delta R.T.: -0.005 min
Response: 1261238
Conc: 50.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.800 min
Delta R.T.: -0.003 min
Response: 831144
Conc: 56.86 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.371 min
Delta R.T.: -0.009 min
Response: 1023892
Conc: 52.49 ng/ml



#16 AR-1242-1


R.T.: 6.176 min
Delta R.T.: 0.000 min
Response: 277594
Conc: 468.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.303 min
Delta R.T.: -0.005 min
Response: 410133
Conc: 511.78 ng/ml




#17 AR-1242-2

R.T.: 6.198 min
Delta R.T.: -0.001 min
Response: 418585
Conc: 469.17 ng/ml



#17 AR-1242-2


R.T.: 5.323 min
Delta R.T.: -0.005 min
Response: 565226
Conc: 499.19 ng/ml



#18 AR-1242-3


R.T.: 6.265 min
Delta R.T.: -0.002 min
Response: 263568
Conc: 475.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.517 min
Delta R.T.: -0.006 min
Response: 315159
Conc: 494.28 ng/ml




#19 AR-1242-4

R.T.: 6.369 min
Delta R.T.: -0.002 min
Response: 211169
Conc: 464.13 ng/ml



#19 AR-1242-4


R.T.: 5.612 min
Delta R.T.: -0.006 min
Response: 295847
Conc: 481.63 ng/ml



#20 AR-1242-5

R.T.: 7.148 min
Delta R.T.: -0.001 min
Response: 208747
Conc: 465.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 6.179 min
Delta R.T.: -0.006 min
Response: 369655
Conc: 513.72 ng/ml