

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011822\
 Data File : PP043068.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Jan 2022 17:07
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 19 04:42:39 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 13 19:45:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.872	4.038	1157448	1269958	48.917	50.847
2)	SA Decachlor...	10.798	9.370	817910	1024295	55.950	52.510
Target Compounds							
26)	L6 AR-1254-1	7.081	6.178	350876	753131	447.164	485.669
27)	L6 AR-1254-2	7.309	6.338	546265	646001	434.029	475.857
28)	L6 AR-1254-3	7.687	6.762	584676	1022453	450.652	478.020
29)	L6 AR-1254-4	7.981	7.001	409879	616608	468.953	494.237
30)	L6 AR-1254-5	8.409	7.432	442685	827109	472.492	485.060



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011822\
Data File : PP043068.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Jan 2022 17:07
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 19 04:42:39 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 13 19:45:07 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.872 min
Delta R.T.: -0.003 min
Response: 1157448
Conc: 48.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.038 min
Delta R.T.: -0.006 min
Response: 1269958
Conc: 50.85 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.798 min
Delta R.T.: -0.005 min
Response: 817910
Conc: 55.95 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.370 min
Delta R.T.: -0.010 min
Response: 1024295
Conc: 52.51 ng/ml



#26 AR-1254-1


R.T.: 7.081 min
Delta R.T.: 0.000 min
Response: 350876
Conc: 447.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.178 min
Delta R.T.: -0.007 min
Response: 753131
Conc: 485.67 ng/ml




#27 AR-1254-2

R.T.: 7.309 min
Delta R.T.: -0.002 min
Response: 546265
Conc: 434.03 ng/ml



#27 AR-1254-2


R.T.: 6.338 min
Delta R.T.: -0.007 min
Response: 646001
Conc: 475.86 ng/ml



#28 AR-1254-3


R.T.: 7.687 min
Delta R.T.: -0.002 min
Response: 584676
Conc: 450.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.762 min
Delta R.T.: -0.007 min
Response: 1022453
Conc: 478.02 ng/ml




#29 AR-1254-4

R.T.: 7.981 min
Delta R.T.: 0.000 min
Response: 409879
Conc: 468.95 ng/ml



#29 AR-1254-4


R.T.: 7.001 min
Delta R.T.: -0.007 min
Response: 616608
Conc: 494.24 ng/ml



#30 AR-1254-5

R.T.: 8.409 min
Delta R.T.: 0.000 min
Response: 442685
Conc: 472.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.432 min
Delta R.T.: -0.007 min
Response: 827109
Conc: 485.06 ng/ml