

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011824\
 Data File : PP062990.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jan 2024 12:53
 Operator : YP\AJ
 Sample : PB158470BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB158470BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 21:55:30 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 11 02:58:31 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.406	3.683	45672537	62913083	17.612	18.031
2) SA Decachlor...	10.180	8.793	32618658	44144330	20.626	18.919

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011824\
Data File : PP062990.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jan 2024 12:53
Operator : YP\AJ
Sample : PB158470BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB158470BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 21:55:30 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011024.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 11 02:58:31 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.406 min
Delta R.T.: 0.010 min
Response: 45672537
Conc: 17.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB158470BL




#1 Tetrachloro-m-xylene

R.T.: 3.683 min
Delta R.T.: 0.001 min
Response: 62913083
Conc: 18.03 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.180 min
Delta R.T.: 0.018 min
Response: 32618658
Conc: 20.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.793 min
Delta R.T.: 0.005 min
Response: 44144330
Conc: 18.92 ng/ml