

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012020\
 Data File : PP030609.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jan 2020 12:55
 Operator : DD\AJ
 Sample : L1161-07
 Misc : LOD 100 PPB
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LOD-MDL-WATER-01-QT1-2020

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 20 16:32:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011720.M
 Quant Title : 8080.M
 QLast Update : Sat Jan 18 00:30:15 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.239	11.407	125.9E6	60341190	342.049	320.261
Target Compounds							
1) T	Dalapon	2.723	2.537	60774517	26631516	106.784	105.995
2) T	3,5-DICHL...	9.653	9.527	69583331	29094468	115.024	115.693
3) T	4-Nitroph...	10.814	10.576	44452870	9849886	100.168	99.397
5) T	DICAMBA	11.548	11.701	168.1E6	78321885	113.223	105.004
6) T	MCPD	11.889	11.901	8826300	1920611	9.383	5.126 #
7) T	MCPA	12.140	12.289	15517514	3841757	11.516	6.423 #
8) T	DICHLORPROP	12.744	12.823	43354900	21887111	119.934	119.237
9) T	2,4-D	13.105	13.313	45543827	23989373	114.637	117.193
10) T	Pentachlo...	13.470	13.874	612.7E6	309.0E6	107.380	105.225
11) T	2,4,5-TP ...	14.353	14.470	226.7E6	115.8E6	114.307	108.860
12) T	2,4,5-T	14.753	15.018	222.8E6	98644627	115.841	107.309
13) T	2,4-DB	15.500	15.704	32113679	16010040	122.479	121.122
14) T	DINOSEB	16.966	16.107	140.0E6	72715655	124.891	125.614
15) T	Picloram	16.750	17.439	264.0E6	143.7E6	102.920	97.617
16) T	DCPA	17.320	17.326	289.2E6	159.0E6	111.397	109.467

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012020\
Data File : PP030609.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jan 2020 12:55
Operator : DD\AJ
Sample : L1161-07
Misc : LOD 100 PPB
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT1-2020

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 20 16:32:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011720.M
Quant Title : 8080.M
QLast Update : Sat Jan 18 00:30:15 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

