

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012022\
 Data File : PP043150.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jan 2022 16:14
 Operator : AJ\MA
 Sample : PB142156BS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 00:24:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 13 19:45:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.871	4.038	442380	489490	18.696	19.598
2)	SA Decachlor...	10.793	9.366	340541	411370	23.295	21.089
Target Compounds							
3)	L1 AR-1016-1	6.173	5.301	300931	426377	405.360	416.479
4)	L1 AR-1016-2	6.196	5.321	450384	597579	391.660	418.429
5)	L1 AR-1016-3	6.262	5.515	281266	323207	395.152	404.099
6)	L1 AR-1016-4	6.367	5.566	225722	252864	372.256	406.856
7)	L1 AR-1016-5	6.682	5.797	204289	292418	356.859	392.175
8)	L2 AR-1221-1	5.107	4.296	48651	39953	182.751	119.650 #
9)	L2 AR-1221-2	5.211	4.397	44379	54264	219.598	223.029
10)	L2 AR-1221-3	5.296	4.486	181197	228297	275.980	291.220
11)	L3 AR-1232-1	5.296	4.486	181197	228297	317.688	350.630
12)	L3 AR-1232-2	5.881	5.321	234767	597579	865.553	952.897
13)	L3 AR-1232-3	6.196	5.515	450384	323207	922.724	977.382
14)	L3 AR-1232-4	6.367	5.610	225722	306938	944.689	1039.857
15)	L3 AR-1232-5	6.468	5.797	165792	292418	1055.227	948.143
16)	L4 AR-1242-1	6.173	5.301	300931	426377	507.928	532.050
17)	L4 AR-1242-2	6.196	5.321	450384	597579	504.816	527.759
18)	L4 AR-1242-3	6.262	5.515	281266	323207	506.921	506.906
19)	L4 AR-1242-4	6.367	5.610	225722	306938	496.113	499.690
20)	L4 AR-1242-5	7.146	6.177	32497	259522	72.447	360.665 #
21)	L5 AR-1248-1	6.173	5.301	300931	426377	734.218	759.034
22)	L5 AR-1248-2	6.468	5.566	165792	252864	294.484	311.547
23)	L5 AR-1248-3	6.682	5.610	204289	306938	301.312	357.118
24)	L5 AR-1248-4	7.104	5.797	28508	292418	39.522	304.688 #
25)	L5 AR-1248-5	7.146	6.220	32497	42693	44.539	44.719
26)	L6 AR-1254-1	7.078	6.177	166490	259522	212.178	167.357
27)	L6 AR-1254-2	7.306	6.336	159407	244648	126.655	180.212 #
28)	L6 AR-1254-3	7.685	6.777	82965	417796	63.947	195.329 #
29)	L6 AR-1254-4	7.978	6.999	53473	45875	61.179	36.771 #
30)	L6 AR-1254-5	8.405	7.429	523266	760047	558.498	445.731
31)	L7 AR-1260-1	7.854	6.896	372369	577565	390.503	446.833
32)	L7 AR-1260-2	8.118	7.092	424688	679081	411.204	434.764
33)	L7 AR-1260-3	8.484	7.249	285376	625850	378.236	433.031
34)	L7 AR-1260-4	8.712	7.734	351842	437057	413.061	385.771
35)	L7 AR-1260-5	9.028	7.982	644668	959157	391.814	382.434
36)	L8 AR-1262-1	8.484	7.429	285376	760047	270.254	839.593 #
37)	L8 AR-1262-2	9.028	7.734	644668	437057	372.430	292.176
38)	L8 AR-1262-3	9.349	8.269	388853	177588	421.876	159.175 #
39)	L8 AR-1262-4	9.423	8.332	99205	693195	183.907	340.042 #
40)	L8 AR-1262-5	10.064	8.832	172161	190498	283.322	205.435 #
41)	L9 AR-1268-1	9.349	8.269	388853	177588	188.615	57.618 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012022\
 Data File : PP043150.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jan 2022 16:14
 Operator : AJ\MA
 Sample : PB142156BS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 00:24:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 13 19:45:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.423	8.332	99205	693195	51.941	243.723 #
43) L9	AR-1268-3	9.649	8.543	11828	19961	7.021	8.376
44) L9	AR-1268-4	10.064	8.832	172161	190498	272.450	186.808 #
45) L9	AR-1268-5	10.466	9.118	60434	64196	11.340	9.037



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012022\
Data File : PP043150.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jan 2022 16:14
Operator : AJ\MA
Sample : PB142156BS
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 00:24:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 13 19:45:07 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.871 min
Delta R.T.: -0.004 min
Response: 442380
Conc: 18.70 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.038 min
Delta R.T.: -0.006 min
Response: 489490
Conc: 19.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.793 min
Delta R.T.: -0.010 min
Response: 340541
Conc: 23.30 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.366 min
Delta R.T.: -0.014 min
Response: 411370
Conc: 21.09 ng/ml



#3 AR-1016-1


R.T.: 6.173 min
Delta R.T.: -0.004 min
Response: 300931
Conc: 405.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.301 min
Delta R.T.: -0.008 min
Response: 426377
Conc: 416.48 ng/ml




#4 AR-1016-2

R.T.: 6.196 min
Delta R.T.: -0.005 min
Response: 450384
Conc: 391.66 ng/ml



#4 AR-1016-2


R.T.: 5.321 min
Delta R.T.: -0.008 min
Response: 597579
Conc: 418.43 ng/ml



#5 AR-1016-3


R.T.: 6.262 min
Delta R.T.: -0.004 min
Response: 281266
Conc: 395.15 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.515 min
Delta R.T.: -0.008 min
Response: 323207
Conc: 404.10 ng/ml




#6 AR-1016-4

R.T.: 6.367 min
Delta R.T.: -0.004 min
Response: 225722
Conc: 372.26 ng/ml



#6 AR-1016-4


R.T.: 5.566 min
Delta R.T.: -0.008 min
Response: 252864
Conc: 406.86 ng/ml



#7 AR-1016-5


R.T.: 6.682 min
Delta R.T.: -0.004 min
Response: 204289
Conc: 356.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.797 min
Delta R.T.: -0.008 min
Response: 292418
Conc: 392.17 ng/ml




#8 AR-1221-1

R.T.: 5.107 min
Delta R.T.: -0.010 min
Response: 48651
Conc: 182.75 ng/ml



#8 AR-1221-1


R.T.: 4.296 min
Delta R.T.: -0.007 min
Response: 39953
Conc: 119.65 ng/ml



#9 AR-1221-2


R.T.: 5.211 min
Delta R.T.: -0.003 min
Response: 44379
Conc: 219.60 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.397 min
Delta R.T.: -0.006 min
Response: 54264
Conc: 223.03 ng/ml




#10 AR-1221-3

R.T.: 5.296 min
Delta R.T.: -0.003 min
Response: 181197
Conc: 275.98 ng/ml



#10 AR-1221-3


R.T.: 4.486 min
Delta R.T.: -0.006 min
Response: 228297
Conc: 291.22 ng/ml



#11 AR-1232-1


R.T.: 5.296 min
Delta R.T.: -0.004 min
Response: 181197
Conc: 317.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.486 min
Delta R.T.: -0.006 min
Response: 228297
Conc: 350.63 ng/ml




#12 AR-1232-2

R.T.: 5.881 min
Delta R.T.: -0.005 min
Response: 234767
Conc: 865.55 ng/ml



#12 AR-1232-2


R.T.: 5.321 min
Delta R.T.: -0.008 min
Response: 597579
Conc: 952.90 ng/ml



#13 AR-1232-3


R.T.: 6.196 min
Delta R.T.: -0.004 min
Response: 450384
Conc: 922.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.515 min
Delta R.T.: -0.008 min
Response: 323207
Conc: 977.38 ng/ml




#14 AR-1232-4

R.T.: 6.367 min
Delta R.T.: -0.005 min
Response: 225722
Conc: 944.69 ng/ml



#14 AR-1232-4


R.T.: 5.610 min
Delta R.T.: -0.009 min
Response: 306938
Conc: 1039.86 ng/ml



#15 AR-1232-5


R.T.: 6.468 min
Delta R.T.: -0.004 min
Response: 165792
Conc: 1055.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.797 min
Delta R.T.: -0.009 min
Response: 292418
Conc: 948.14 ng/ml




#16 AR-1242-1

R.T.: 6.173 min
Delta R.T.: -0.003 min
Response: 300931
Conc: 507.93 ng/ml



#16 AR-1242-1


R.T.: 5.301 min
Delta R.T.: -0.008 min
Response: 426377
Conc: 532.05 ng/ml



#17 AR-1242-2


R.T.: 6.196 min
Delta R.T.: -0.004 min
Response: 450384
Conc: 504.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.321 min
Delta R.T.: -0.008 min
Response: 597579
Conc: 527.76 ng/ml




#18 AR-1242-3

R.T.: 6.262 min
Delta R.T.: -0.004 min
Response: 281266
Conc: 506.92 ng/ml



#18 AR-1242-3


R.T.: 5.515 min
Delta R.T.: -0.008 min
Response: 323207
Conc: 506.91 ng/ml



#19 AR-1242-4


R.T.: 6.367 min
Delta R.T.: -0.004 min
Response: 225722
Conc: 496.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.610 min
Delta R.T.: -0.008 min
Response: 306938
Conc: 499.69 ng/ml




#20 AR-1242-5

R.T.: 7.146 min
Delta R.T.: -0.003 min
Response: 32497
Conc: 72.45 ng/ml



#20 AR-1242-5


R.T.: 6.177 min
Delta R.T.: -0.008 min
Response: 259522
Conc: 360.67 ng/ml



#21 AR-1248-1


R.T.: 6.173 min
Delta R.T.: -0.003 min
Response: 300931
Conc: 734.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.301 min
Delta R.T.: -0.007 min
Response: 426377
Conc: 759.03 ng/ml




#22 AR-1248-2

R.T.: 6.468 min
Delta R.T.: -0.003 min
Response: 165792
Conc: 294.48 ng/ml



#22 AR-1248-2


R.T.: 5.566 min
Delta R.T.: -0.007 min
Response: 252864
Conc: 311.55 ng/ml



#23 AR-1248-3


R.T.: 6.682 min
Delta R.T.: -0.003 min
Response: 204289
Conc: 301.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.610 min
Delta R.T.: -0.008 min
Response: 306938
Conc: 357.12 ng/ml




#24 AR-1248-4

R.T.: 7.104 min
Delta R.T.: -0.006 min
Response: 28508
Conc: 39.52 ng/ml



#24 AR-1248-4


R.T.: 5.797 min
Delta R.T.: -0.008 min
Response: 292418
Conc: 304.69 ng/ml



#25 AR-1248-5


R.T.: 7.146 min
Delta R.T.: -0.003 min
Response: 32497
Conc: 44.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.220 min
Delta R.T.: -0.009 min
Response: 42693
Conc: 44.72 ng/ml




#26 AR-1254-1

R.T.: 7.078 min
Delta R.T.: -0.003 min
Response: 166490
Conc: 212.18 ng/ml



#26 AR-1254-1


R.T.: 6.177 min
Delta R.T.: -0.009 min
Response: 259522
Conc: 167.36 ng/ml



#27 AR-1254-2


R.T.: 7.306 min
Delta R.T.: -0.004 min
Response: 159407
Conc: 126.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.336 min
Delta R.T.: -0.008 min
Response: 244648
Conc: 180.21 ng/ml




#28 AR-1254-3

R.T.: 7.685 min
Delta R.T.: -0.004 min
Response: 82965
Conc: 63.95 ng/ml



#28 AR-1254-3


R.T.: 6.777 min
Delta R.T.: 0.008 min
Response: 417796
Conc: 195.33 ng/ml



#29 AR-1254-4


R.T.: 7.978 min
Delta R.T.: -0.004 min
Response: 53473
Conc: 61.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.999 min
Delta R.T.: -0.009 min
Response: 45875
Conc: 36.77 ng/ml




#30 AR-1254-5

R.T.: 8.405 min
Delta R.T.: -0.005 min
Response: 523266
Conc: 558.50 ng/ml



#30 AR-1254-5


R.T.: 7.429 min
Delta R.T.: -0.010 min
Response: 760047
Conc: 445.73 ng/ml



#31 AR-1260-1


R.T.: 7.854 min
Delta R.T.: -0.005 min
Response: 372369
Conc: 390.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.896 min
Delta R.T.: -0.010 min
Response: 577565
Conc: 446.83 ng/ml




#32 AR-1260-2

R.T.: 8.118 min
Delta R.T.: -0.005 min
Response: 424688
Conc: 411.20 ng/ml



#32 AR-1260-2


R.T.: 7.092 min
Delta R.T.: -0.010 min
Response: 679081
Conc: 434.76 ng/ml



#33 AR-1260-3


R.T.: 8.484 min
Delta R.T.: -0.006 min
Response: 285376
Conc: 378.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.249 min
Delta R.T.: -0.011 min
Response: 625850
Conc: 433.03 ng/ml




#34 AR-1260-4

R.T.: 8.712 min
Delta R.T.: -0.006 min
Response: 351842
Conc: 413.06 ng/ml



#34 AR-1260-4


R.T.: 7.734 min
Delta R.T.: -0.011 min
Response: 437057
Conc: 385.77 ng/ml



#35 AR-1260-5


R.T.: 9.028 min
Delta R.T.: -0.006 min
Response: 644668
Conc: 391.81 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.982 min
Delta R.T.: -0.010 min
Response: 959157
Conc: 382.43 ng/ml




#36 AR-1262-1

R.T.: 8.484 min
Delta R.T.: -0.006 min
Response: 285376
Conc: 270.25 ng/ml



#36 AR-1262-1


R.T.: 7.429 min
Delta R.T.: -0.010 min
Response: 760047
Conc: 839.59 ng/ml



#37 AR-1262-2


R.T.: 9.028 min
Delta R.T.: -0.006 min
Response: 644668
Conc: 372.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.734 min
Delta R.T.: -0.010 min
Response: 437057
Conc: 292.18 ng/ml




#38 AR-1262-3

R.T.: 9.349 min
Delta R.T.: 0.005 min
Response: 388853
Conc: 421.88 ng/ml



#38 AR-1262-3


R.T.: 8.269 min
Delta R.T.: -0.011 min
Response: 177588
Conc: 159.17 ng/ml



#39 AR-1262-4


R.T.: 9.423 min
Delta R.T.: -0.006 min
Response: 99205
Conc: 183.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.332 min
Delta R.T.: -0.012 min
Response: 693195
Conc: 340.04 ng/ml




#40 AR-1262-5

R.T.: 10.064 min
Delta R.T.: -0.009 min
Response: 172161
Conc: 283.32 ng/ml



#40 AR-1262-5


R.T.: 8.832 min
Delta R.T.: -0.011 min
Response: 190498
Conc: 205.44 ng/ml



#41 AR-1268-1


R.T.: 9.349 min
Delta R.T.: 0.008 min
Response: 388853
Conc: 188.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.269 min
Delta R.T.: -0.011 min
Response: 177588
Conc: 57.62 ng/ml




#42 AR-1268-2

R.T.: 9.423 min
Delta R.T.: -0.011 min
Response: 99205
Conc: 51.94 ng/ml



#42 AR-1268-2


R.T.: 8.332 min
Delta R.T.: -0.013 min
Response: 693195
Conc: 243.72 ng/ml



#43 AR-1268-3


R.T.: 9.649 min
Delta R.T.: -0.008 min
Response: 11828
Conc: 7.02 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.543 min
Delta R.T.: -0.011 min
Response: 19961
Conc: 8.38 ng/ml




#44 AR-1268-4

R.T.: 10.064 min
Delta R.T.: -0.009 min
Response: 172161
Conc: 272.45 ng/ml



#44 AR-1268-4


R.T.: 8.832 min
Delta R.T.: -0.012 min
Response: 190498
Conc: 186.81 ng/ml



#45 AR-1268-5

R.T.: 10.466 min
Delta R.T.: -0.010 min
Response: 60434
Conc: 11.34 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.118 min
Delta R.T.: -0.013 min
Response: 64196
Conc: 9.04 ng/ml