

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012122\
 Data File : PP043180.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Jan 2022 16:37
 Operator : AJ\MA
 Sample : N1221-02 10X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 GJA261L-END-1-2

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/25/2022
 Supervised By :mohammad ahmed 01/25/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 23:23:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 13 19:45:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.869	4.037	32872	43317	1.389m	1.734
2) SA Decachlor...	10.792	9.364	16098	36125	1.101m	1.852 #
Target Compounds						
31) L7 AR-1260-1	7.848	6.893	5475	19561	5.742m	15.133 #
32) L7 AR-1260-2	8.118	7.092	8179	20824	7.920	13.332m#
33) L7 AR-1260-3	8.484	7.246	7653	16204	10.143m	11.211
34) L7 AR-1260-4	8.710	7.731	6357	11724	7.463m	10.348m#
35) L7 AR-1260-5	9.027	7.979	24534	30720	14.911	12.249m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012122\
Data File : PP043180.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Jan 2022 16:37
Operator : AJ\MA
Sample : N1221-02 10X
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
GJA261L-END-1-2


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/25/2022
Supervised By :mohammad ahmed 01/25/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 23:23:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 13 19:45:07 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.869 min
Delta R.T.: -0.006 min
Response: 32872
Conc: 1.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
GJA261L-END-1-2


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/25/2022
Supervised By :mohammad ahmed 01/25/2022




#1 Tetrachloro-m-xylene

R.T.: 4.037 min
Delta R.T.: -0.007 min
Response: 43317
Conc: 1.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.792 min
Delta R.T.: -0.010 min
Response: 16098
Conc: 1.10 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.364 min
Delta R.T.: -0.006 min
Response: 36125
Conc: 1.85 ng/ml




#31 AR-1260-1

R.T.: 7.848 min
Delta R.T.: -0.011 min
Response: 5475
Conc: 5.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
GJA261L-END-1-2


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/25/2022
Supervised By :mohammad ahmed 01/25/2022




#31 AR-1260-1

R.T.: 6.893 min
Delta R.T.: -0.012 min
Response: 19561
Conc: 15.13 ng/ml




#32 AR-1260-2

R.T.: 8.118 min
Delta R.T.: -0.004 min
Response: 8179
Conc: 7.92 ng/ml



#32 AR-1260-2

R.T.: 7.092 min
Delta R.T.: -0.010 min
Response: 20824
Conc: 13.33 ng/ml m




#33 AR-1260-3

R.T.: 8.484 min
Delta R.T.: -0.006 min
Response: 7653
Conc: 10.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
GJA261L-END-1-2


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/25/2022
Supervised By :mohammad ahmed 01/25/2022




#33 AR-1260-3

R.T.: 7.246 min
Delta R.T.: -0.013 min
Response: 16204
Conc: 11.21 ng/ml




#34 AR-1260-4

R.T.: 8.710 min
Delta R.T.: -0.008 min
Response: 6357
Conc: 7.46 ng/ml m



#34 AR-1260-4

R.T.: 7.731 min
Delta R.T.: -0.014 min
Response: 11724
Conc: 10.35 ng/ml m




#35 AR-1260-5

R.T.: 9.027 min
Delta R.T.: -0.007 min
Response: 24534
Conc: 14.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
GJA261L-END-1-2

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/25/2022
Supervised By :mohammad ahmed 01/25/2022



#35 AR-1260-5

R.T.: 7.979 min
Delta R.T.: -0.013 min
Response: 30720
Conc: 12.25 ng/ml m