

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012125\
 Data File : PP069246.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2025 08:21
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 09:47:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 07 11:05:55 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.672	3.982	71762345	46951939	51.580	48.098
2) SA Decachlor...	10.510	9.098	53859489	53277244	48.241	44.349
Target Compounds						
21) L5 AR-1248-1	5.831	5.080	14491605	11182935	492.449	486.232
22) L5 AR-1248-2	6.103	5.320	21338884	15808031	494.862	463.013
23) L5 AR-1248-3	6.308	5.363	23680811	16292870	501.951	461.125
24) L5 AR-1248-4	6.707	5.537	26463370	18952200	492.793	460.432
25) L5 AR-1248-5	6.747	5.933	26236928	18557511	504.389	479.693



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012125\
Data File : PP069246.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2025 08:21
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 09:47:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 07 11:05:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.672 min
Delta R.T.: 0.000 min
Response: 71762345
Conc: 51.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.982 min
Delta R.T.: -0.001 min
Response: 46951939
Conc: 48.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.510 min
Delta R.T.: -0.014 min
Response: 53859489
Conc: 48.24 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.098 min
Delta R.T.: -0.013 min
Response: 53277244
Conc: 44.35 ng/ml



#21 AR-1248-1


R.T.: 5.831 min
Delta R.T.: 0.000 min
Response: 14491605
Conc: 492.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.080 min
Delta R.T.: -0.005 min
Response: 11182935
Conc: 486.23 ng/ml




#22 AR-1248-2

R.T.: 6.103 min
Delta R.T.: -0.001 min
Response: 21338884
Conc: 494.86 ng/ml



#22 AR-1248-2


R.T.: 5.320 min
Delta R.T.: -0.006 min
Response: 15808031
Conc: 463.01 ng/ml



#23 AR-1248-3


R.T.: 6.308 min
Delta R.T.: -0.001 min
Response: 23680811
Conc: 501.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.363 min
Delta R.T.: -0.006 min
Response: 16292870
Conc: 461.13 ng/ml




#24 AR-1248-4

R.T.: 6.707 min
Delta R.T.: -0.001 min
Response: 26463370
Conc: 492.79 ng/ml



#24 AR-1248-4


R.T.: 5.537 min
Delta R.T.: -0.006 min
Response: 18952200
Conc: 460.43 ng/ml



#25 AR-1248-5

R.T.: 6.747 min
Delta R.T.: -0.002 min
Response: 26236928
Conc: 504.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.933 min
Delta R.T.: -0.007 min
Response: 18557511
Conc: 479.69 ng/ml