

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP012219.M
 Title : 8080.M
 Last Update : Wed Jan 23 06:13:00 2019
 Response Via : Initial Calibration

Calibration Files

750 =PP028936.D 1000 =PP028937.D 1500 =PP028938.D
 500 =PP028935.D 200 =PP028934.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	5.962	6.132	6.021	5.902	5.871	5.978 E5	1.74
2) T	3,5-DICHLOROBENZOI...	6.695	5.477	5.256	5.292	5.334	5.611 E5	10.91
3) T	4-Nitrophenol	6.639	5.443	5.376	4.992	3.822	5.255 E5	19.24
4) S	2,4-DCAA	2.500	2.519	2.455	2.412	2.341	2.446 E5	2.93
5) T	DICAMBA	1.521	1.547	1.509	1.498	1.515	1.518 E6	1.20
6) T	MCPP	4.989	5.136	5.282	4.363	3.307	4.615 E5	17.57
7) T	MCPA	7.697	8.201	7.843	6.948	4.726	7.083 E5	19.69
8) T	DICHLORPROP	3.309	3.326	3.076	3.081	3.385	3.235 E5	4.51
9) T	2,4-D	3.510	3.704	3.602	3.598	2.908	3.464 E5	9.19
10) T	Pentachlorophenol	7.864	7.834	7.269	7.764	7.892	7.724 E6	3.35
11) T	2,4,5-TP (SILVEX)	2.724	2.762	2.655	2.730	2.754	2.725 E6	1.55
12) T	2,4,5-T	2.662	2.676	2.555	2.727	2.906	2.705 E6	4.75
13) T	2,4-DB	1.991	2.056	2.055	1.804	1.425	1.866 E5	14.33
14) T	DINOSEB	2.385	2.338	2.214	2.442	2.599	2.395 E6	5.91
15) T	Picloram	4.180	4.345	4.313	3.978	3.564	4.076 E6	7.87
16) T	DCPA	4.175	4.165	3.971	4.140	4.126	4.115 E6	2.02

Signal #2 Calibration Files

750 =PP028936.D 1000 =PP028937.D 1500 =PP028938.D
 500 =PP028935.D 200 =PP028934.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	5.810	5.896	5.873	5.727	5.589	5.779 E5	2.16
2) T	3,5-DICHLOROBENZOI...	5.854	5.709	5.506	5.525	5.502	5.619 E5	2.79
3) T	4-Nitrophenol	2.777	2.331	2.300	2.037	1.621	2.213 E5	19.18
4) S	2,4-DCAA	4.392	4.510	4.382	4.339	4.065	4.338 E5	3.81
5) T	DICAMBA	1.679	1.716	1.651	1.659	1.630	1.667 E6	1.95
6) T	MCPP	9.434	9.661	9.800	8.684	6.587	8.833 E5	15.03
7) T	MCPA	1.510	1.580	1.583	1.353	1.077	1.421 E6	15.01
8) T	DICHLORPROP	4.326	4.419	4.261	4.250	4.193	4.290 E5	2.01
9) T	2,4-D	5.132	5.274	5.111	5.026	4.746	5.058 E5	3.87
10) T	Pentachlorophenol	6.654	6.538	5.958	6.706	6.931	6.558 E6	5.55
11) T	2,4,5-TP (SILVEX)	2.538	2.562	2.428	2.552	2.529	2.522 E6	2.14
12) T	2,4,5-T	2.378	2.403	2.264	2.457	2.454	2.391 E6	3.30
13) T	2,4-DB	2.782	2.906	2.865	2.621	2.387	2.712 E5	7.81
14) T	DINOSEB	1.884	1.906	1.815	1.846	1.887	1.868 E6	1.95
15) T	Picloram	3.731	3.855	3.721	3.709	3.473	3.698 E6	3.75
16) T	DCPA	3.271	3.245	3.014	3.283	3.329	3.229 E6	3.83

(#) = Out of Range