

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012219\
 Data File : PP028937.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Jan 2019 10:57
 Operator : EI\MA
 Sample : HSTDICC1000
 Misc : hexane
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDICC1000

Manual Integrations
 APPROVED

Sohil
 1/23/2019 10:02:54 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 23 06:01:31 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012219.M
 Quant Title : 8080.M
 QLast Update : Wed Jan 23 05:57:57 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.069	12.183	251.9E6	451.0E6	1007.733	1026.731
Target Compounds							
1) T	Dalapon	3.327	2.999	558.0E6	536.6E6	935.908	923.498
2) T	3,5-DICHL...	10.483	10.275	509.3E6	531.0E6	760.783	906.996
3) T	4-Nitroph...	11.648	11.329	495.3E6	212.1E6	745.993	763.976
5) T	DICAMBA	12.390	12.492	1454.6E6	1612.6E6	956.602	960.598
6) T	MCPD	12.735	12.687	48276099	90816280	96.774	96.262
7) T	MCPA	12.977	13.076	76273026	146.9E6	99.100m	97.332
8) T	DICHLORPROP	13.587	13.607	312.6E6	415.4E6	944.735	960.218
9) T	2,4-D	13.940	14.097	348.2E6	495.7E6	991.851	966.054
10) T	Pentachlo...	14.356	14.705	7441.9E6	6211.4E6	946.379	933.496
11) T	2,4,5-TP ...	15.220	15.276	2624.2E6	2433.5E6	963.425	958.750
12) T	2,4,5-T	15.611	15.918	2542.7E6	2282.8E6	955.311	960.150
13) T	2,4-DB	16.375	16.520	195.3E6	276.1E6	981.233	992.281
14) T	DINOSEB	17.860	16.910	2197.4E6	1791.7E6	921.535	951.000
15) T	Picloram	17.629	18.292	4128.0E6	3662.1E6	987.601	981.483
16) T	DCPA	18.224	18.168	3998.3E6	3115.5E6	957.624	952.394

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012219\
Data File : PP028937.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Jan 2019 10:57
Operator : EI\MA
Sample : HSTDICC1000
Misc : hexane
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
Client Sampled :
HSTDICC1000

Manual Integrations
APPROVED

Sohil
1/23/2019 10:02:54 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 23 06:01:31 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012219.M
Quant Title : 8080.M
QLast Update : Wed Jan 23 05:57:57 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

