

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012221\
 Data File : PP032905.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Jan 2021 19:31
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 23 00:40:21 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 14 08:12:55 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.737	4.006	2883027	2379558	49.638	50.586
2)	SA Decachlor...	10.572	9.282	2172724	1934356	55.071	51.635
Target Compounds							
3)	L1 AR-1016-1	6.038	5.254	801414	682260	493.304	502.285
4)	L1 AR-1016-2	6.061	5.274	1184978	990313	479.555	488.247
5)	L1 AR-1016-3	6.126	5.465	722519	549436	486.364	502.860
6)	L1 AR-1016-4	6.232	5.516	616804	416326	495.567	514.285
7)	L1 AR-1016-5	6.546	5.744	575314	524956	506.945	487.223
31)	L7 AR-1260-1	7.715	6.836	941701	938587	493.978	492.023
32)	L7 AR-1260-2	7.979	7.034	1226849	1129564	507.974	495.671
33)	L7 AR-1260-3	8.344	7.188	1064250	1088086	528.005	497.588
34)	L7 AR-1260-4	8.574	7.669	1051945	932353	510.541	510.232
35)	L7 AR-1260-5	8.886	7.918	2406235	2203048	531.186	512.137



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP012221\
Data File : PP032905.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Jan 2021 19:31
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 23 00:40:21 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011421.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 14 08:12:55 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.737 min
Delta R.T.: -0.004 min
Response: 2883027
Conc: 49.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.006 min
Delta R.T.: -0.007 min
Response: 2379558
Conc: 50.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.572 min
Delta R.T.: -0.005 min
Response: 2172724
Conc: 55.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.282 min
Delta R.T.: -0.014 min
Response: 1934356
Conc: 51.63 ng/ml



#3 AR-1016-1


R.T.: 6.038 min
Delta R.T.: -0.005 min
Response: 801414
Conc: 493.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.254 min
Delta R.T.: -0.009 min
Response: 682260
Conc: 502.29 ng/ml




#4 AR-1016-2

R.T.: 6.061 min
Delta R.T.: -0.005 min
Response: 1184978
Conc: 479.55 ng/ml



#4 AR-1016-2


R.T.: 5.274 min
Delta R.T.: -0.010 min
Response: 990313
Conc: 488.25 ng/ml



#5 AR-1016-3


R.T.: 6.126 min
Delta R.T.: -0.005 min
Response: 722519
Conc: 486.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.465 min
Delta R.T.: -0.010 min
Response: 549436
Conc: 502.86 ng/ml




#6 AR-1016-4

R.T.: 6.232 min
Delta R.T.: -0.006 min
Response: 616804
Conc: 495.57 ng/ml



#6 AR-1016-4


R.T.: 5.516 min
Delta R.T.: -0.010 min
Response: 416326
Conc: 514.28 ng/ml



#7 AR-1016-5


R.T.: 6.546 min
Delta R.T.: -0.006 min
Response: 575314
Conc: 506.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.744 min
Delta R.T.: -0.010 min
Response: 524956
Conc: 487.22 ng/ml




#31 AR-1260-1

R.T.: 7.715 min
Delta R.T.: -0.006 min
Response: 941701
Conc: 493.98 ng/ml



#31 AR-1260-1


R.T.: 6.836 min
Delta R.T.: -0.010 min
Response: 938587
Conc: 492.02 ng/ml



#32 AR-1260-2


R.T.: 7.979 min
Delta R.T.: -0.006 min
Response: 1226849
Conc: 507.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 7.034 min
Delta R.T.: -0.010 min
Response: 1129564
Conc: 495.67 ng/ml




#33 AR-1260-3

R.T.: 8.344 min
Delta R.T.: -0.006 min
Response: 1064250
Conc: 528.01 ng/ml



#33 AR-1260-3


R.T.: 7.188 min
Delta R.T.: -0.011 min
Response: 1088086
Conc: 497.59 ng/ml



#34 AR-1260-4


R.T.: 8.574 min
Delta R.T.: -0.005 min
Response: 1051945
Conc: 510.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.669 min
Delta R.T.: -0.011 min
Response: 932353
Conc: 510.23 ng/ml



#35 AR-1260-5

R.T.: 8.886 min
Delta R.T.: -0.005 min
Response: 2406235
Conc: 531.19 ng/ml



#35 AR-1260-5

R.T.: 7.918 min
Delta R.T.: -0.011 min
Response: 2203048
Conc: 512.14 ng/ml